

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071118\
 Data File : PS000969.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2018 05:21
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 7/12/2018 4:01:52 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 07:31:44 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 09 02:00:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.633	11.047	149.3E6	146.8E6	723.614	673.890m
Target Compounds							
1) T	Dalapon	3.019	2.429	167.3E6	162.5E6	622.967	603.229
2) T	3,5-DICHL...	10.048	9.193	199.7E6	186.6E6	666.805	634.469
3) T	4-Nitroph...	11.225	10.193	81896309	73989568	597.087m	613.182
5) T	DICAMBA	11.950	11.335	544.0E6	514.2E6	690.860	655.107
6) T	MCPD	12.289	11.558	39824562	35629866	72.501m	68.140m
7) T	MCPA	12.540	11.937	58699073	49662154	64.318	63.891
8) T	DICHLORPROP	13.142	12.459	137.6E6	129.4E6	682.789	640.623
9) T	2,4-D	13.500	12.931	184.7E6	178.7E6	678.510	645.246
10) T	Pentachlo...	13.890	13.487	2151.0E6	2101.8E6	708.250	669.666
11) T	2,4,5-TP ...	14.759	14.101	855.3E6	828.8E6	700.108	660.440
12) T	2,4,5-T	15.156	14.672	842.6E6	664.6E6	691.967	642.305
13) T	2,4-DB	15.907	15.318	111.4E6	102.3E6	687.855	668.130
14) T	DINOSEB	17.410	15.749	567.6E6	531.8E6	656.302	635.023
15) T	Picloram	17.168	16.932	1195.0E6	1016.4E6	659.817	667.538
16) T	DCPA	17.749	16.988	1033.1E6	1119.4E6	709.629	651.148

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071118\
Data File : PS000969.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2018 05:21
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
7/12/2018 4:01:52 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 07:31:44 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
Quant Title : 8080.M
QLast Update : Mon Jul 09 02:00:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

