

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071218\
 Data File : PS000998.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2018 00:27
 Operator : UA
 Sample : J3947-08MSD
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 PAR-4MSD

Manual Integrations
 APPROVED

Sohil
 7/13/2018 1:02:24 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 13 01:38:12 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 09 02:00:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.662	11.078	46728506	42087409	226.486m	193.241
Target Compounds							
1) T	Dalapon	3.048	2.459	83439998	83267549	310.754	309.054
2) T	3,5-DICHL...	10.078	9.224	106.7E6	106.4E6	356.153	361.695
5) T	DICAMBA	11.978	11.366	327.8E6	308.3E6	416.277m	392.777m
6) T	MCP	12.316	11.588	20731740	18379967	37.742	35.151
7) T	MCPA	12.565	11.963	34239293	28491575	37.517m	36.655m
8) T	DICHLORPROP	13.170	12.490	90916484	86811202	451.052m	429.806m
9) T	2,4-D	13.528	12.961	118.2E6	119.3E6	434.075	430.891
10) T	Pentachlo...	13.918	13.517	777.0E6	741.1E6	255.834	236.140
11) T	2,4,5-TP ...	14.787	14.131	530.6E6	513.3E6	434.266	408.999
12) T	2,4,5-T	15.185	14.701	521.9E6	331.3E6	428.559	320.156 #
13) T	2,4-DB	15.935	15.348	68910097	62217424	425.675	406.172
14) T	DINOSEB	17.436	15.779	154.2E6	139.7E6	178.338	166.796
15) T	Picloram	17.194	16.962	619.5E6	448.1E6	342.040	294.313
16) T	DCPA	17.776	17.018	455.3E6	569.7E6	312.770	331.390

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071218\
Data File : PS000998.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2018 00:27
Operator : UA
Sample : J3947-08MSD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
PAR-4MSD

Manual Integrations
APPROVED

Sohil
7/13/2018 1:02:24 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 13 01:38:12 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
Quant Title : 8080.M
QLast Update : Mon Jul 09 02:00:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

