

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071224\
 Data File : PS027148.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jul 2024 10:07
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 12 12:50:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071224.M
 Quant Title : 8080.M
 QLast Update : Fri Jul 12 12:49:33 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.205	7.752	1211.4E6	1142.0E6	479.034	492.520
Target Compounds							
1) T	Dalapon	2.612	2.690	222217	7352675	<MDL	1.772 #
2) T	3,5-DICHL...	6.342f	0.000	24320	0	<MDL	N.D. #
3) T	4-Nitroph...	7.004	0.000	3590519	0	2.240	N.D. #
5) T	DICAMBA	7.365	7.942	312251	13025620	<MDL	1.236 #
6) T	MCPD	7.565	0.000	2903639	0	<MDL	N.D. #
7) T	MCPA	7.716	0.000	372018	0	<MDL	N.D. #
8) T	DICHLORPROP	8.097	8.699f	107883	1789428	<MDL	<MDL #
9) T	2,4-D	0.000	8.944f	0	780173	N.D.	<MDL
10) T	Pentachlo...	8.611	0.000	1534304	0	<MDL	N.D. #
11) T	2,4,5-TP ...	9.184	0.000	132580	0	<MDL	N.D. #
12) T	2,4,5-T	9.491	0.000	656093	0	<MDL	N.D. #
13) T	2,4-DB	10.071f	10.918f	2344701	-815596	1.085	N.D. #
14) T	DINOSEB	11.242	0.000	1258448	0	<MDL	N.D. #
15) T	Picloram	11.065	12.359	9110718	8371245	<MDL	<MDL #
16) T	DCPA	0.000	12.305	0	1051754	N.D.	<MDL



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071224\
Data File : PS027148.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2024 10:07
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 12:50:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071224.M
Quant Title : 8080.M
QLast Update : Fri Jul 12 12:49:33 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Dalapon


R.T.: 2.612 min
Delta R.T.: -0.007 min
Response: 222217
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.690 min
Delta R.T.: -0.010 min
Response: 7352675
Conc: 1.77 ng/ml




#3 4-Nitrophenol

R.T.: 7.004 min
Delta R.T.: 0.000 min
Response: 3590519
Conc: 2.24 ng/ml



#3 4-Nitrophenol


R.T.: 0.000 min
Exp R.T. : 7.280 min
Response: 0
Conc: N.D.



#4 2,4-DCAA


R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 1211364780
Conc: 479.03 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 7.752 min
Delta R.T.: 0.000 min
Response: 1142017963
Conc: 492.52 ng/ml




#5 DICAMBA

R.T.: 7.365 min
Delta R.T.: -0.024 min
Response: 312251
Conc: N.D.



#5 DICAMBA


R.T.: 7.942 min
Delta R.T.: -0.009 min
Response: 13025620
Conc: 1.24 ng/ml



#13 2,4-DB

R.T.: 10.071 min
Delta R.T.: 0.032 min
Response: 2344701
Conc: 1.09 ng/ml

Instrument :
ECD_S
ClientSampleId :



#13 2,4-DB

R.T.: 10.918 min
Delta R.T.: 0.038 min
Response: -815596
Conc: N.D.