

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071224\
Data File : PS027161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2024 15:22
Operator : AR\AJ
Sample : P3193-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
FILL-MATERIAL-1

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/15/2024
Supervised By :Ankita Jodhani 07/15/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 23:59:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071224.M
Quant Title : 8080.M
QLast Update : Fri Jul 12 12:49:33 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
4) S 2,4-DCAA	7.203	7.755	973.1E6	650.3E6	384.809m	280.450 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071224\
Data File : PS027161.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jul 2024 15:22
Operator : AR\AJ
Sample : P3193-02
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_S
ClientSampleId :
FILL-MATERIAL-1


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/15/2024
Supervised By : Ankita Jodhani 07/15/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 12 23:59:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071224.M
Quant Title : 8080.M
QLast Update : Fri Jul 12 12:49:33 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#4 2,4-DCAA

R.T.: 7.203 min
Delta R.T.: -0.002 min
Response: 973091504
Conc: 384.81 ng/ml

Instrument :
ECD_S
ClientSampleId :
FILL-MATERIAL-1

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 07/15/2024
Supervised By :Ankita Jodhani 07/15/2024



#4 2,4-DCAA

R.T.: 7.755 min
Delta R.T.: 0.002 min
Response: 650285833
Conc: 280.45 ng/ml