

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071318\
 Data File : PS001025.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2018 17:45
 Operator : UA
 Sample : J3930-04MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 NYS-RBR-200030MS

Manual Integrations
 APPROVED

Sohil
 7/16/2018 4:56:24 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 00:42:34 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 09 02:00:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.649	11.064	52197484	49052222	252.993m	225.220
Target Compounds							
1) T	Dalapon	3.035	2.446	101.6E6	100.1E6	378.488m	371.671m
2) T	3,5-DICHL...	10.065	9.210	121.5E6	113.2E6	405.732m	384.898m
5) T	DICAMBA	11.966	11.351	376.7E6	360.2E6	478.455	458.834
6) T	MCPD	12.304	11.575	27739415	21494645	50.500m	41.107
7) T	MCPA	12.555	11.950	37323557	33228836	40.896	42.749m
8) T	DICHLORPROP	13.158	12.477	100.2E6	98392587	496.957	487.146
9) T	2,4-D	13.515	12.948	134.2E6	135.1E6	492.884	487.878
10) T	Pentachlo...	13.905	13.503	936.2E6	908.0E6	308.270m	289.294
11) T	2,4,5-TP ...	14.775	14.117	584.9E6	575.6E6	478.724m	458.652m
12) T	2,4,5-T	15.171	14.688	569.5E6	344.7E6	467.717m	333.181m#
13) T	2,4-DB	15.923	15.334	78071122	71561245	482.265	467.170
14) T	DINOSEB	17.424	15.766	185.3E6	172.9E6	214.214	206.489
15) T	Picloram	17.182	16.949	675.0E6	513.9E6	372.724	337.502
16) T	DCPA	17.764	17.005	513.5E6	633.0E6	352.713	368.222

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071318\
Data File : PS001025.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2018 17:45
Operator : UA
Sample : J3930-04MS
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
NYS-RBR-200030MS

Manual Integrations
APPROVED

Sohil
7/16/2018 4:56:24 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 00:42:34 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
Quant Title : 8080.M
QLast Update : Mon Jul 09 02:00:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

