

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071320\
 Data File : PS011180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2020 19:19
 Operator : DD\AJ
 Sample : L3216-07
 Misc : 100 PPB LOD-WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2020

Manual Integrations
 APPROVED

Ankita
 7/14/2020 11:12:00 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 01:55:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.226	6.918	482.3E6	314.5E6	497.019	558.511
Target Compounds							
1) T	Dalapon	3.030	2.611	112.5E6	87682667	93.492	114.219m
2) T	3,5-DICHL...	6.439	6.069	129.0E6	77461527	96.431	92.233
3) T	4-Nitroph...	7.060	6.528	33726283	27800613	76.504	84.789
5) T	DICAMBA	7.417	7.086	350.5E6	216.3E6	90.438	94.458m
6) T	MCPD	7.581	7.176	29413800	16620536	9.713	8.711
7) T	MCPA	7.730	7.387	40256186	46252462	9.439	16.047 #
8) T	DICHLORPROP	8.110	7.713	94051652	63137697	100.319	104.892
9) T	2,4-D	8.341	8.009	95114950	87268958	100.007	116.411
10) T	Pentachlo...	8.650	8.436	1446.7E6	882.4E6	109.443	126.603
11) T	2,4,5-TP ...	9.200	8.797	552.1E6	341.1E6	104.254	111.569
12) T	2,4,5-T	9.487	9.178	492.6E6	327.4E6	97.618	107.962
13) T	2,4-DB	10.040	9.685	51359852	42867853	79.686	100.276 #
14) T	DINOSEB	11.199	10.016	340.1E6	212.5E6	88.238	99.094
15) T	Picloram	11.027	11.005	438.9E6	372.3E6	69.559	85.127
16) T	DCPA	11.466	10.945	551.1E6	442.3E6	85.536	116.958 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071320\
 Data File : PS011180.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2020 19:19
 Operator : DD\AJ
 Sample : L3216-07
 Misc : 100 PPB LOD-WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2020

Manual Integrations
 APPROVED

Ankita
 7/14/2020 11:12:00 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 01:55:13 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

