

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071321\
 Data File : PS016061.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2021 13:45
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 06:21:49 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 30 00:31:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.187	6.765	1154.9E6	388.0E6	752.259	755.963
Target Compounds							
1) T	Dalapon	3.017	2.518	1500.4E6	487.3E6	691.783	697.158
2) T	3,5-DICHL...	6.405	5.957	1530.1E6	549.6E6	705.453	713.984
3) T	4-Nitroph...	7.014	6.386	590.2E6	216.9E6	667.355	687.154
5) T	DICAMBA	7.378	6.929	4153.5E6	1531.5E6	724.980	727.068
6) T	MCPD	7.548	7.026	323.5E6	121.5E6	73.782	73.240
7) T	MCPA	7.697	7.238	457.7E6	182.7E6	71.979	69.330
8) T	DICHLORPROP	8.071	7.549	978.6E6	375.4E6	703.556	712.962
9) T	2,4-D	8.299	7.840	1111.8E6	450.4E6	693.525	714.164
10) T	Pentachlo...	8.608	8.267	13762.3E6	5115.9E6	694.473	711.209
11) T	2,4,5-TP ...	9.159	8.628	5241.8E6	2059.6E6	706.408	729.953
12) T	2,4,5-T	9.442	9.005	5564.0E6	2089.9E6	712.542	724.459
13) T	2,4-DB	9.992	9.511	1138.5E6	273.3E6	685.376	726.878
14) T	DINOSEB	11.152	9.838	3818.4E6	1509.6E6	696.501	739.336
15) T	Picloram	10.970	10.816	7703.3E6	3064.6E6	700.288	736.585
16) T	DCPA	11.418	10.769	7655.9E6	3171.5E6	709.744	737.481

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071321\
Data File : PS016061.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2021 13:45
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 06:21:49 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
Quant Title : 8080.M
QLast Update : Wed Jun 30 00:31:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

