

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071321\
 Data File : PS016066.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2021 18:58
 Operator : DD\AJ
 Sample : M2967-01MSD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 1-5ft-GRACE-AVE-TP-1MSD

Manual Integrations
 APPROVED

mohammad
 7/14/2021 12:21:10 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 06:25:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 30 00:31:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.186	6.765	746.4E6	219.8E6	486.195	428.163m
Target Compounds							
1) T	Dalapon	3.014	2.519	580.8E6	224.9E6	267.789	321.847m
2) T	3,5-DICHL...	6.404	5.957	519.7E6	169.7E6	239.625	220.421m
3) T	4-Nitroph...	7.015	6.388	28664647	7811883	32.414m	24.750
5) T	DICAMBA	7.377	6.929	1553.3E6	568.7E6	271.122	269.999
6) T	MCPD	7.544	7.020	95178422	36003251	21.710	21.701
7) T	MCPA	7.691	7.231	157.3E6	78461102	24.735	29.767
8) T	DICHLORPROP	8.069	7.549	323.3E6	124.6E6	232.407m	236.603m
9) T	2,4-D	8.298	7.839	537.0E6	197.0E6	334.954	312.369
10) T	Pentachlo...	8.607	8.266	2098.4E6	796.0E6	105.892	110.656
11) T	2,4,5-TP ...	9.157	8.628	1335.6E6	551.9E6	179.996	195.584
12) T	2,4,5-T	9.441	9.004	1631.7E6	624.2E6	208.958	216.374
13) T	2,4-DB	9.993	9.510	148.9E6	71939469	89.630	191.360 #
14) T	DINOSEB	11.114	9.839	398.2E6	67463621	72.640	33.041m#
15) T	Picloram	10.968	10.814	3455.7E6	1373.2E6	314.149	330.048
16) T	DCPA	11.418	10.767	2805.3E6	1330.3E6	260.072	309.333

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071321\
Data File : PS016066.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2021 18:58
Operator : DD\AJ
Sample : M2967-01MSD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
1-5ft-GRACE-AVE-TP-1MSD

Manual Integrations
APPROVED

mohammad
7/14/2021 12:21:10 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 06:25:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
Quant Title : 8080.M
QLast Update : Wed Jun 30 00:31:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

