

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071422\
 Data File : PS019961.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2022 14:18
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 21:46:14 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062722.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 27 17:26:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.193	6.778	2020.5E6	545.7E6	779.036	671.885
Target Compounds							
1) T	Dalapon	3.015	2.529	2742.9E6	614.1E6	565.876	578.575
2) T	3,5-DICHL...	6.408	5.960	2478.1E6	804.8E6	723.440	629.772
3) T	4-Nitroph...	7.028	6.399	796.8E6	272.8E6	748.651	630.795
5) T	DICAMBA	7.381	6.941	7633.9E6	2350.1E6	743.969	644.212
6) T	MCPD	7.549	7.037	568.4E6	180.5E6	70.841	62.868
7) T	MCPA	7.697	7.249	821.6E6	252.3E6	70.113	62.325
8) T	DICHLORPROP	8.071	7.561	1915.8E6	603.0E6	707.579	641.028
9) T	2,4-D	8.303	7.854	2105.3E6	711.4E6	708.831	631.388
10) T	Pentachlo...	8.606	8.277	26717.0E6	8254.2E6	739.816	668.512
11) T	2,4,5-TP ...	9.156	8.638	11535.6E6	3505.6E6	721.173	661.490
12) T	2,4,5-T	9.443	9.016	12190.2E6	3214.3E6	662.508	657.948
13) T	2,4-DB	9.997	9.522	2434.3E6	412.3E6	608.328	634.764
14) T	DINOSEB	11.144	9.845	10758.5E6	2407.2E6	661.748	675.380
15) T	Picloram	10.975	10.829	19483.4E6	3969.4E6	624.812	610.346
16) T	DCPA	11.413	10.773	16927.2E6	4297.4E6	656.166	663.094

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071422\
Data File : PS019961.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2022 14:18
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 21:46:14 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062722.M
Quant Title : 8080.M
QLast Update : Mon Jun 27 17:26:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

