

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071520\
 Data File : PS011233.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2020 12:41
 Operator : DD\AJ
 Sample : L3287-03MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 FK-GF-03-012-BMSD

Manual Integrations
 APPROVED

Ankita
 7/16/2020 10:39:57 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 17:21:00 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.226	6.917	223.8E6	98452595	230.693	174.864m
Target Compounds							
1) T	Dalapon	3.034	2.617	204.8E6	160.2E6	170.208	208.685m
2) T	3,5-DICHL...	6.439	6.068	168.9E6	104.4E6	126.192	124.287m
3) T	4-Nitroph...	7.057	6.525	48214941	30330228	109.370	92.504m
5) T	DICAMBA	7.418	7.087	584.3E6	348.6E6	150.749	152.240
6) T	MCPD	7.582	7.178	37653946	16914202	12.435	8.864 #
7) T	MCPA	7.730	7.391	64036706	87049920	15.015	30.201 #
8) T	DICHLORPROP	8.110	7.713	123.5E6	96445116	131.699	160.227
9) T	2,4-D	8.339	8.006	151.6E6	147.9E6	159.400	197.312
10) T	Pentachlo...	8.650	8.436	1677.9E6	1073.3E6	126.937	154.001
11) T	2,4,5-TP ...	9.199	8.796	614.0E6	391.0E6	115.936	127.884m
12) T	2,4,5-T	9.484	9.176	603.0E6	409.4E6	119.511	134.967
13) T	2,4-DB	10.037	9.681	64467157	73939294	100.022	172.958 #
14) T	DINOSEB	11.196	10.016	122.0E6	76829996	31.639m	35.819m
15) T	Picloram	11.017	10.995	1838.5E6	1142.5E6	291.372	261.239
16) T	DCPA	11.464	10.943	1940.7E6	1315.2E6	301.221	347.760

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071520\
 Data File : PS011233.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2020 12:41
 Operator : DD\AJ
 Sample : L3287-03MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 FK-GF-03-012-BMSD

Manual Integrations
 APPROVED

Ankita
 7/16/2020 10:39:57 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 17:21:00 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

