

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071522\
 Data File : PS019975.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2022 14:41
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/18/2022
 Supervised By : Ankita Jodhani 07/18/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 01:36:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062722.M
 Quant Title : 8080.M
 QLast Update : Mon Jun 27 17:26:48 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.192	6.777	1791.9E6	542.6E6	690.860	668.047m
Target Compounds							
1) T	Dalapon	3.015	2.528	2579.4E6	638.5E6	532.150	601.615
2) T	3,5-DICHL...	6.408	5.960	2222.8E6	822.9E6	648.929	643.965
3) T	4-Nitroph...	7.028	6.400	681.4E6	274.1E6	640.157	633.706
5) T	DICAMBA	7.380	6.941	6812.0E6	2410.9E6	663.868	660.897
6) T	MCPD	7.549	7.037	498.5E6	185.5E6	62.130	64.633
7) T	MCPA	7.696	7.249	757.5E6	257.5E6	64.648	63.624
8) T	DICHLORPROP	8.071	7.561	1792.6E6	614.6E6	662.091	653.293
9) T	2,4-D	8.303	7.854	1925.7E6	730.5E6	648.340	648.320
10) T	Pentachlo...	8.606	8.277	24854.3E6	8445.5E6	688.236	684.000
11) T	2,4,5-TP ...	9.156	8.638	10663.2E6	3593.9E6	666.630	678.146
12) T	2,4,5-T	9.442	9.016	11955.2E6	3297.3E6	649.737	674.941
13) T	2,4-DB	9.996	9.522	2519.9E6	424.7E6	629.719	653.863
14) T	DINOSEB	11.144	9.844	10835.4E6	2485.9E6	666.477	697.466
15) T	Picloram	10.974	10.828	20414.0E6	4282.1E6	654.655	658.418
16) T	DCPA	11.412	10.772	17033.9E6	4397.7E6	660.300	678.575

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071522\
Data File : PS019975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2022 14:41
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/18/2022
Supervised By : Ankita Jodhani 07/18/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 01:36:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS062722.M
Quant Title : 8080.M
QLast Update : Mon Jun 27 17:26:48 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

