

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071619\
 Data File : PS005514.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jul 2019 21:59
 Operator : SM\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Ankita
 7/17/2019 9:54:47 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 01:27:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 20 11:04:04 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	5.596	5.312	1134.0E6	768.8E6	772.562	766.947
Target Compounds							
1) T	Dalapon	2.129	1.827	1297.7E6	897.0E6	704.032	680.048
2) T	3,5-DICHL...	4.947	4.605	1449.6E6	911.1E6	721.276	690.931
3) T	4-Nitroph...	5.443	4.988	465.1E6	339.5E6	642.111	677.005
5) T	DICAMBA	5.755	5.445	4382.8E6	2765.5E6	768.275	738.775
6) T	MCPD	5.915	5.535	346.5E6	209.2E6	79.152	75.106
7) T	MCPA	6.052	5.720	470.8E6	304.7E6	74.971	75.220
8) T	DICHLORPROP	6.394	5.999	1085.5E6	704.5E6	759.034	729.600
9) T	2,4-D	6.610	6.272	1136.3E6	788.3E6	730.900	706.786
10) T	Pentachlo...	6.880	6.648	14643.6E6	9163.3E6	744.230	712.750
11) T	2,4,5-TP ...	7.416	7.011	6131.7E6	3849.6E6	751.594	722.889
12) T	2,4,5-T	7.690	7.372	6207.8E6	3580.6E6	771.359	701.956
13) T	2,4-DB	8.222	7.861	801.5E6	481.2E6	690.067	737.084
14) T	DINOSEB	9.322	8.175	4815.2E6	2827.8E6	753.358	699.386
15) T	Picloram	9.150	9.116	7153.7E6	4399.1E6	622.542	582.938m
16) T	DCPA	9.593	9.075	7484.8E6	4913.2E6	750.342	721.740

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071619\
Data File : PS005514.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2019 21:59
Operator : SM\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Ankita
7/17/2019 9:54:47 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 01:27:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS061919.M
Quant Title : 8080.M
QLast Update : Thu Jun 20 11:04:04 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

