

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071620\
 Data File : PS011238.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jul 2020 12:59
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 16 13:34:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.227	6.918	762.0E6	443.8E6	785.333	788.325
Target Compounds							
1) T	Dalapon	3.038	2.624	800.3E6	497.8E6	665.054	648.482
2) T	3,5-DICHL...	6.440	6.068	979.1E6	617.8E6	731.661	735.605
3) T	4-Nitroph...	7.057	6.524	315.8E6	235.2E6	716.330	717.476
5) T	DICAMBA	7.418	7.087	2885.0E6	1711.4E6	744.360	747.333
6) T	MCPD	7.588	7.185	227.2E6	144.3E6	75.045	75.649
7) T	MCPA	7.738	7.401	316.2E6	211.0E6	74.136	73.209
8) T	DICHLORPROP	8.110	7.712	687.3E6	443.9E6	733.047	737.504
9) T	2,4-D	8.339	8.006	707.6E6	551.2E6	743.985	735.238
10) T	Pentachlo...	8.650	8.437	9774.3E6	5046.1E6	739.430	724.010
11) T	2,4,5-TP ...	9.199	8.796	3928.3E6	2302.3E6	741.785	753.070
12) T	2,4,5-T	9.484	9.175	3751.3E6	2290.1E6	743.446	755.061
13) T	2,4-DB	10.037	9.682	478.4E6	319.5E6	742.217	747.471
14) T	DINOSEB	11.198	10.016	2827.0E6	1630.2E6	733.406	760.006
15) T	Picloram	11.019	10.998	4480.2E6	3121.8E6	710.031	713.826
16) T	DCPA	11.463	10.943	4734.7E6	2852.1E6	734.866	754.132

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071620\
Data File : PS011238.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2020 12:59
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 16 13:34:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
Quant Title : 8080.M
QLast Update : Mon Jul 13 16:10:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

