

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071621\
 Data File : PS016119.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jul 2021 10:58
 Operator : DD\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HEXANE

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 00:31:54 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 30 00:31:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
----------	------	------	--------	--------	-------	-------

System Monitoring Compounds

Target Compounds

5) T	DICAMBA	0.000	6.986f	0 91063401	N.D.	43.232
6) T	MCP	0.000	6.986f	0 91063401	N.D.	54.888
9) T	2,4-D	8.266	0.000	15746448	0	9.823 N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071621\
Data File : PS016119.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2021 10:58
Operator : DD\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HEXANE

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 00:31:54 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
Quant Title : 8080.M
QLast Update : Wed Jun 30 00:31:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#4 2,4-DCAA


R.T.: 0.000 min
Exp R.T. : 7.188 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
HEXANE




#4 2,4-DCAA

R.T.: 0.000 min
Exp R.T. : 6.767 min
Response: 0
Conc: N.D.




#5 DICAMBA

R.T.: 0.000 min
Exp R.T. : 7.380 min
Response: 0
Conc: N.D.



#5 DICAMBA


R.T.: 6.986 min
Delta R.T.: 0.056 min
Response: 91063401
Conc: 43.23 ng/ml



#6 MCPP


R.T.: 0.000 min
Exp R.T. : 7.550 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
HEXANE




#6 MCPP

R.T.: 6.986 min
Delta R.T.: -0.042 min
Response: 91063401
Conc: 54.89 ug/ml



#9 2,4-D

R.T.: 8.266 min
Delta R.T.: -0.033 min
Response: 15746448
Conc: 9.82 ng/ml



#9 2,4-D

R.T.: 0.000 min
Exp R.T. : 7.841 min
Response: 0
Conc: N.D.