

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071720\
 Data File : PS011278.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jul 2020 11:23
 Operator : DD\AJ
 Sample : L3187-08MSD
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SU-03-071420MSD

Manual Integrations
 APPROVED

Sohil
 7/20/2020 3:04:28 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 14:21:42 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.226	6.917	354.1E6	227.9E6	364.971	404.709
Target Compounds							
1) T	Dalapon	3.036	2.620	567.6E6	511.0E6	471.635	665.681m#
2) T	3,5-DICHL...	6.439	6.068	656.6E6	401.4E6	490.680	477.992
5) T	DICAMBA	7.417	7.086	2202.7E6	1305.1E6	568.303	569.901
6) T	MCP	7.585	7.183	390.0E6	118.3E6	128.800m	62.008 #
7) T	MCPA	7.735	7.398	283.3E6	219.9E6	66.431	76.304
8) T	DICHLORPROP	8.109	7.711	546.0E6	370.2E6	582.421	614.945
9) T	2,4-D	8.338	8.004	679.2E6	503.3E6	714.117	671.344m
10) T	Pentachlo...	8.649	8.436	4481.2E6	2575.0E6	339.001	369.459
11) T	2,4,5-TP ...	9.198	8.786	3374.5E6	4223.3E6	637.200	1381.390m#
12) T	2,4,5-T	9.483	9.175	3176.0E6	1965.2E6	629.424	647.950
13) T	2,4-DB	10.035	9.681	370.1E6	266.7E6	574.250	623.848
14) T	DINOSEB	11.196	10.015	1096.9E6	577.4E6	284.572	269.210
15) T	Picloram	11.016	10.996	3883.6E6	2265.0E6	615.478	517.909
16) T	DCPA	11.462	10.943	3985.9E6	2451.6E6	618.653	648.237

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071720\
Data File : PS011278.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jul 2020 11:23
Operator : DD\AJ
Sample : L3187-08MSD
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SU-03-071420MSD

Manual Integrations
APPROVED

Sohil
7/20/2020 3:04:28 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 14:21:42 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
Quant Title : 8080.M
QLast Update : Mon Jul 13 16:10:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

