

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071724\
 Data File : PS027189.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jul 2024 15:03
 Operator : AR\AJ
 Sample : P3247-27
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-CTW2-BL-06

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 18 03:18:52 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071224.M
 Quant Title : 8080.M
 QLast Update : Thu Jun 27 04:33:52 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.204	7.755	595.0E6	480.3E6	235.288	207.142
Target Compounds							
1) T	Dalapon	2.579f	2.672f	411.7E6	489.9E6	123.844	118.045
3) T	4-Nitroph...	6.993	0.000	24686847	0	15.399	N.D. #
5) T	DICAMBA	0.000	7.942	0	9130813	N.D.	<MDL
6) T	MCP	7.557	8.051	8371367	1224475	1.091	<MDL #
7) T	MCPA	7.708	0.000	109.5E6	0	10.745	N.D. #
8) T	DICHLORPROP	8.117f	8.646	3121584	874187	1.163	<MDL #
9) T	2,4-D	0.000	8.993	0	3442620	N.D.	1.194
10) T	Pentachlo...	8.603	9.525	4793438	1489009	<MDL	<MDL #
11) T	2,4,5-TP ...	9.182	0.000	2832799	0	<MDL	N.D. #
12) T	2,4,5-T	9.467	10.293	179740	50189255	<MDL	3.141 #
13) T	2,4-DB	10.045	10.924f	47060156	740008	21.778	<MDL #
14) T	DINOSEB	11.268f	11.270	6646986	29854182	<MDL	2.672 #
15) T	Picloram	11.033	12.357	51381979	3751915	2.638	<MDL #
16) T	DCPA	11.500f	12.303	64909688	6570150	3.616	<MDL #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071724\
Data File : PS027189.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jul 2024 15:03
Operator : AR\AJ
Sample : P3247-27
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
WC-CTW2-BL-06

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 03:18:52 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071224.M
Quant Title : 8080.M
QLast Update : Thu Jun 27 04:33:52 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.579 min
Delta R.T.: -0.040 min
Response: 411732760
Conc: 123.84 ng/ml

Instrument :
ECD_S
ClientSampleId :
WC-CTW2-BL-06




#1 Dalapon

R.T.: 2.672 min
Delta R.T.: -0.028 min
Response: 489898392
Conc: 118.04 ng/ml




#3 4-Nitrophenol

R.T.: 6.993 min
Delta R.T.: -0.012 min
Response: 24686847
Conc: 15.40 ng/ml



#3 4-Nitrophenol


R.T.: 0.000 min
Exp R.T. : 7.280 min
Response: 0
Conc: N.D.



#4 2,4-DCAA


R.T.: 7.204 min
Delta R.T.: -0.001 min
Response: 594987547
Conc: 235.29 ng/ml

Instrument :
ECD_S
ClientSampleId :
WC-CTW2-BL-06




#4 2,4-DCAA

R.T.: 7.755 min
Delta R.T.: 0.003 min
Response: 480305410
Conc: 207.14 ng/ml




#6 MCP

R.T.: 7.557 min
Delta R.T.: -0.008 min
Response: 8371367
Conc: 1.09 ug/ml



#6 MCP


R.T.: 8.051 min
Delta R.T.: 0.002 min
Response: 1224475
Conc: N.D.



#7 MCPA


R.T.: 7.708 min
Delta R.T.: -0.003 min
Response: 109488624
Conc: 10.74 ug/ml

Instrument :
ECD_S
ClientSampleId :
WC-CTW2-BL-06




#7 MCPA

R.T.: 0.000 min
Exp R.T. : 8.292 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 8.117 min
Delta R.T.: 0.032 min
Response: 3121584
Conc: 1.16 ng/ml



#8 DICHLORPROP


R.T.: 8.646 min
Delta R.T.: -0.019 min
Response: 874187
Conc: N.D.



#9 2,4-D


R.T.: 0.000 min
Exp R.T.: 8.314 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :
WC-CTW2-BL-06




#9 2,4-D

R.T.: 8.993 min
Delta R.T.: 0.000 min
Response: 3442620
Conc: 1.19 ng/ml




#12 2,4,5-T

R.T.: 9.467 min
Delta R.T.: -0.002 min
Response: 179740
Conc: N.D.



#12 2,4,5-T


R.T.: 10.293 min
Delta R.T.: -0.022 min
Response: 50189255
Conc: 3.14 ng/ml



#13 2,4-DB


R.T.: 10.045 min
Delta R.T.: 0.006 min
Response: 47060156
Conc: 21.78 ng/ml

Instrument :
ECD_S
ClientSampleId :
WC-CTW2-BL-06




#13 2,4-DB

R.T.: 10.924 min
Delta R.T.: 0.044 min
Response: 740008
Conc: N.D.




#14 DINOSEB

R.T.: 11.268 min
Delta R.T.: 0.033 min
Response: 6646986
Conc: N.D.



#14 DINOSEB


R.T.: 11.270 min
Delta R.T.: 0.012 min
Response: 29854182
Conc: 2.67 ng/ml



#15 Picloram


R.T.: 11.033 min
Delta R.T.: -0.017 min
Response: 51381979
Conc: 2.64 ng/ml

Instrument :
ECD_S
ClientSampleId :
WC-CTW2-BL-06




#15 Picloram

R.T.: 12.357 min
Delta R.T.: 0.005 min
Response: 3751915
Conc: N.D.



#16 DCPA

R.T.: 11.500 min
Delta R.T.: -0.031 min
Response: 64909688
Conc: 3.62 ng/ml



#16 DCPA

R.T.: 12.303 min
Delta R.T.: 0.004 min
Response: 6570150
Conc: N.D.