

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071818\
 Data File : PS001121.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Jul 2018 11:16
 Operator : UA
 Sample : J3819-02
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 18 11:42:26 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 09 02:00:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.649	11.055	51044599	43664305	247.405	200.482
Target Compounds							
1) T	Dalapon	0.000	2.471	0	413482	N.D.	1.535 #
5) T	DICAMBA	12.028	11.299	2486745	5926303	3.158	7.550 #
6) T	MCPD	12.341	0.000	10310601	0	18.771	N.D. #
8) T	DICHLORPROP	0.000	12.423	0	1028884	N.D.	5.094 #
12) T	2,4,5-T	0.000	14.683	0	4306368	N.D.	4.162 #
14) T	DINOSEB	0.000	15.823	0	3595397	N.D.	4.293 #
15) T	Picloram	17.182	16.935	8301418	3413721	4.584	2.242 #
16) T	DCPA	0.000	16.997	0	8703793	N.D.	5.063 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071818\
Data File : PS001121.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Jul 2018 11:16
Operator : UA
Sample : J3819-02
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 18 11:42:26 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
Quant Title : 8080.M
QLast Update : Mon Jul 09 02:00:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 0.000 min
Exp R.T.: 3.046 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.471 min
Delta R.T.: 0.016 min
Response: 413482
Conc: 1.53 ng/ml




#4 2,4-DCAA

R.T.: 11.649 min
Delta R.T.: -0.004 min
Response: 51044599
Conc: 247.40 ng/ml



#4 2,4-DCAA


R.T.: 11.055 min
Delta R.T.: -0.018 min
Response: 43664305
Conc: 200.48 ng/ml



#5 DICAMBA


R.T.: 12.028 min
Delta R.T.: 0.058 min
Response: 2486745
Conc: 3.16 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 11.299 min
Delta R.T.: -0.063 min
Response: 5926303
Conc: 7.55 ng/ml




#6 MCP

R.T.: 12.341 min
Delta R.T.: 0.029 min
Response: 10310601
Conc: 18.77 ug/ml



#6 MCP


R.T.: 0.000 min
Exp R.T. : 11.591 min
Response: 0
Conc: N.D.



#8 DICHLORPROP


R.T.: 0.000 min
Exp R.T.: 13.161 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 12.423 min
Delta R.T.: -0.063 min
Response: 1028884
Conc: 5.09 ng/ml




#12 2,4,5-T

R.T.: 0.000 min
Exp R.T.: 15.174 min
Response: 0
Conc: N.D.



#12 2,4,5-T


R.T.: 14.683 min
Delta R.T.: -0.015 min
Response: 4306368
Conc: 4.16 ng/ml



#14 DINOSEB


R.T.: 0.000 min
Exp R.T.: 17.427 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#14 DINOSEB

R.T.: 15.823 min
Delta R.T.: 0.047 min
Response: 3595397
Conc: 4.29 ng/ml




#15 Picloram

R.T.: 17.182 min
Delta R.T.: -0.001 min
Response: 8301418
Conc: 4.58 ng/ml



#15 Picloram


R.T.: 16.935 min
Delta R.T.: -0.024 min
Response: 3413721
Conc: 2.24 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 17.766 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 16.997 min
Delta R.T.: -0.018 min
Response: 8703793
Conc: 5.06 ng/ml