

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071918\
 Data File : PS001155.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jul 2018 17:28
 Operator : UA
 Sample : J4019-03
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 02:30:03 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 09 02:00:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.660	11.076	50084419	42405416	242.751	194.702
Target Compounds							
1) T	Dalapon	3.003	2.449	2119109	1186215	7.892	4.403 #
2) T	3,5-DICHL...	0.000	9.163	0	10722396	N.D.	36.452 #
5) T	DICAMBA	11.910	11.321	3844250	6180648	4.882	7.874 #
6) T	MCP	12.354	0.000	5905735	0	10.751	N.D. #
8) T	DICHLORPROP	0.000	12.440	0	1259177	N.D.	6.234 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071918\
Data File : PS001155.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jul 2018 17:28
Operator : UA
Sample : J4019-03
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 02:30:03 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
Quant Title : 8080.M
QLast Update : Mon Jul 09 02:00:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.003 min
Delta R.T.: -0.043 min
Response: 2119109
Conc: 7.89 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.449 min
Delta R.T.: -0.006 min
Response: 1186215
Conc: 4.40 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 0.000 min
Exp R.T. : 10.070 min
Response: 0
Conc: N.D.



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.163 min
Delta R.T.: -0.056 min
Response: 10722396
Conc: 36.45 ng/ml



#4 2,4-DCAA


R.T.: 11.660 min
Delta R.T.: 0.007 min
Response: 50084419
Conc: 242.75 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 11.076 min
Delta R.T.: 0.003 min
Response: 42405416
Conc: 194.70 ng/ml




#5 DICAMBA

R.T.: 11.910 min
Delta R.T.: -0.060 min
Response: 3844250
Conc: 4.88 ng/ml



#5 DICAMBA


R.T.: 11.321 min
Delta R.T.: -0.041 min
Response: 6180648
Conc: 7.87 ng/ml



#6 MCP


R.T.: 12.354 min
Delta R.T.: 0.042 min
Response: 5905735
Conc: 10.75 ug/ml

Instrument :
ECD_S
ClientSampleId :




#6 MCP

R.T.: 0.000 min
Exp R.T. : 11.591 min
Response: 0
Conc: N.D.



#8 DICHLORPROP

R.T.: 0.000 min
Exp R.T. : 13.161 min
Response: 0
Conc: N.D.



#8 DICHLORPROP

R.T.: 12.440 min
Delta R.T.: -0.045 min
Response: 1259177
Conc: 6.23 ng/ml