

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071918\
 Data File : PS001143.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jul 2018 10:29
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 7/20/2018 1:10:03 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 19 11:02:09 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 09 02:00:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.645	11.060	144.0E6	146.5E6	697.997	672.597
Target Compounds							
1) T	Dalapon	3.037	2.446	169.3E6	162.0E6	630.389	601.360
2) T	3,5-DICHL...	10.062	9.207	192.2E6	178.2E6	641.595	605.857
3) T	4-Nitroph...	11.238	10.206	85126553	75051117	620.638	621.979
5) T	DICAMBA	11.960	11.347	533.7E6	504.7E6	677.787	642.955
6) T	MCPD	12.299	11.571	37683034	33437259	68.602m	63.947m
7) T	MCPA	12.551	11.948	58034344	50047075	63.589	64.386m
8) T	DICHLORPROP	13.152	12.472	132.9E6	126.3E6	659.518	625.527
9) T	2,4-D	13.511	12.943	183.7E6	179.3E6	674.822	647.304
10) T	Pentachlo...	13.900	13.500	2071.1E6	2026.0E6	681.924	645.533
11) T	2,4,5-TP ...	14.769	14.113	834.6E6	808.9E6	683.148	644.579
12) T	2,4,5-T	15.167	14.683	841.3E6	853.9E6	690.844	825.226m
13) T	2,4-DB	15.918	15.329	109.7E6	98911143	677.401	645.718
14) T	DINOSEB	17.419	15.761	569.8E6	528.7E6	658.886	631.244
15) T	Picloram	17.178	16.943	1886.6E6	982.3E6	1041.715m	645.093 #
16) T	DCPA	17.759	17.000	1006.5E6	1818.0E6	691.334	1057.499 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS071918\
Data File : PS001143.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jul 2018 10:29
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
7/20/2018 1:10:03 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 19 11:02:09 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
Quant Title : 8080.M
QLast Update : Mon Jul 09 02:00:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

