

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072018\
 Data File : PS001163.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2018 13:27
 Operator : UA
 Sample : J4016-21RE
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 14:43:33 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 09 02:00:25 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml	

System Monitoring Compounds								
4) S	2,4-DCAA	11.646	11.061	56343181	43401208	273.086	199.274	#
Target Compounds								
1) T	Dalapon	2.985	0.000	2731992	0	10.175	N.D.	#
3) T	4-Nitroph...	11.241	0.000	17237600	0	125.675	N.D.	#
5) T	DICAMBA	12.024	11.304	5836809	6676228	7.413	8.505	
6) T	MCP	12.336	0.000	10591712	0	19.282	N.D.	#
15) T	Picloram	17.177	17.001	5240767	8176600	2.894	5.370	#
16) T	DCPA	0.000	17.001	0	8176600	N.D.	4.756	#



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072018\
Data File : PS001163.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2018 13:27
Operator : UA
Sample : J4016-21RE
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 14:43:33 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS070818.M
Quant Title : 8080.M
QLast Update : Mon Jul 09 02:00:25 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 2.985 min
Delta R.T.: -0.061 min
Response: 2731992
Conc: 10.17 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 0.000 min
Exp R.T. : 2.455 min
Response: 0
Conc: N.D.




#3 4-Nitrophenol

R.T.: 11.241 min
Delta R.T.: -0.005 min
Response: 17237600
Conc: 125.68 ng/ml



#3 4-Nitrophenol


R.T.: 0.000 min
Exp R.T. : 10.219 min
Response: 0
Conc: N.D.



#4 2,4-DCAA


R.T.: 11.646 min
Delta R.T.: -0.007 min
Response: 56343181
Conc: 273.09 ng/ml

Instrument :
ECD_S
ClientSampleId :




#4 2,4-DCAA

R.T.: 11.061 min
Delta R.T.: -0.013 min
Response: 43401208
Conc: 199.27 ng/ml




#5 DICAMBA

R.T.: 12.024 min
Delta R.T.: 0.054 min
Response: 5836809
Conc: 7.41 ng/ml



#5 DICAMBA


R.T.: 11.304 min
Delta R.T.: -0.057 min
Response: 6676228
Conc: 8.51 ng/ml



#6 MCPP


R.T.: 12.336 min
Delta R.T.: 0.024 min
Response: 10591712
Conc: 19.28 ug/ml

Instrument :
ECD_S
ClientSampleId :




#6 MCPP

R.T.: 0.000 min
Exp R.T. : 11.591 min
Response: 0
Conc: N.D.




#15 Picloram

R.T.: 17.177 min
Delta R.T.: -0.006 min
Response: 5240767
Conc: 2.89 ng/ml



#15 Picloram


R.T.: 17.001 min
Delta R.T.: 0.042 min
Response: 8176600
Conc: 5.37 ng/ml



#16 DCPA

R.T.: 0.000 min
Exp R.T.: 17.766 min
Response: 0
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :



#16 DCPA

R.T.: 17.001 min
Delta R.T.: -0.014 min
Response: 8176600
Conc: 4.76 ng/ml