

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072021\
 Data File : PS016154.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2021 12:35
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 03:59:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 30 00:31:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.186	6.765	1166.0E6	401.2E6	759.506	781.611
Target Compounds							
1) T	Dalapon	3.016	2.517	1508.4E6	496.2E6	695.460	709.959
2) T	3,5-DICHL...	6.404	5.957	1549.5E6	571.1E6	714.386	741.931
3) T	4-Nitroph...	7.014	6.386	592.7E6	222.9E6	670.250	706.321
5) T	DICAMBA	7.377	6.928	4190.8E6	1582.3E6	731.478	751.172
6) T	MCPD	7.547	7.026	324.2E6	125.5E6	73.950	75.660
7) T	MCPA	7.696	7.238	458.0E6	189.7E6	72.017	71.962
8) T	DICHLORPROP	8.070	7.548	992.2E6	389.5E6	713.330	739.638
9) T	2,4-D	8.298	7.839	1124.3E6	463.7E6	701.372	735.262
10) T	Pentachlo...	8.607	8.267	14071.2E6	5317.9E6	710.060	739.283
11) T	2,4,5-TP ...	9.157	8.627	5323.8E6	2129.4E6	717.462	754.688
12) T	2,4,5-T	9.441	9.005	5670.8E6	2142.9E6	726.218	742.837
13) T	2,4-DB	9.991	9.510	1139.8E6	281.0E6	686.148	747.390
14) T	DINOSEB	11.150	9.837	3896.3E6	1554.6E6	710.711	761.385
15) T	Picloram	10.968	10.815	7287.6E6	2906.3E6	662.497	698.524
16) T	DCPA	11.417	10.768	7680.3E6	3262.1E6	712.008	758.545



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072021\
Data File : PS016154.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2021 12:35
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 03:59:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
Quant Title : 8080.M
QLast Update : Wed Jun 30 00:31:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.016 min
Delta R.T.: -0.002 min
Response: 1508365972
Conc: 695.46 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.517 min
Delta R.T.: -0.002 min
Response: 496208717
Conc: 709.96 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 6.404 min
Delta R.T.: -0.002 min
Response: 1549474523
Conc: 714.39 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 5.957 min
Delta R.T.: -0.001 min
Response: 571094871
Conc: 741.93 ng/ml



#3 4-Nitrophenol


R.T.: 7.014 min
Delta R.T.: -0.001 min
Response: 592718650
Conc: 670.25 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 6.386 min
Delta R.T.: 0.000 min
Response: 222933051
Conc: 706.32 ng/ml




#4 2,4-DCAA

R.T.: 7.186 min
Delta R.T.: -0.002 min
Response: 1165992913
Conc: 759.51 ng/ml



#4 2,4-DCAA


R.T.: 6.765 min
Delta R.T.: -0.001 min
Response: 401214415
Conc: 781.61 ng/ml



#5 DICAMBA


R.T.: 7.377 min
Delta R.T.: -0.002 min
Response: 4190762903
Conc: 731.48 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 6.928 min
Delta R.T.: -0.002 min
Response: 1582270118
Conc: 751.17 ng/ml




#6 MCP

R.T.: 7.547 min
Delta R.T.: -0.003 min
Response: 324198731
Conc: 73.95 ug/ml



#6 MCP


R.T.: 7.026 min
Delta R.T.: -0.002 min
Response: 125526702
Conc: 75.66 ug/ml



#7 MCPA


R.T.: 7.696 min
Delta R.T.: -0.003 min
Response: 457965364
Conc: 72.02 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 7.238 min
Delta R.T.: -0.002 min
Response: 189679197
Conc: 71.96 ug/ml




#8 DICHLORPROP

R.T.: 8.070 min
Delta R.T.: -0.002 min
Response: 992243779
Conc: 713.33 ng/ml



#8 DICHLORPROP


R.T.: 7.548 min
Delta R.T.: -0.002 min
Response: 389482884
Conc: 739.64 ng/ml



#9 2,4-D


R.T.: 8.298 min
Delta R.T.: -0.002 min
Response: 1124339068
Conc: 701.37 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 7.839 min
Delta R.T.: -0.001 min
Response: 463666876
Conc: 735.26 ng/ml




#10 Pentachlorophenol

R.T.: 8.607 min
Delta R.T.: -0.003 min
Response: 14071161998
Conc: 710.06 ng/ml



#10 Pentachlorophenol


R.T.: 8.267 min
Delta R.T.: -0.003 min
Response: 5317864258
Conc: 739.28 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 9.157 min
Delta R.T.: -0.002 min
Response: 5323811988
Conc: 717.46 ng/ml

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 8.627 min
Delta R.T.: -0.002 min
Response: 2129403277
Conc: 754.69 ng/ml




#12 2,4,5-T

R.T.: 9.441 min
Delta R.T.: -0.002 min
Response: 5670803247
Conc: 726.22 ng/ml



#12 2,4,5-T


R.T.: 9.005 min
Delta R.T.: -0.001 min
Response: 2142885952
Conc: 742.84 ng/ml



#13 2,4-DB


R.T.: 9.991 min
Delta R.T.: 0.000 min
Response: 1139751956
Conc: 686.15 ng/ml

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 9.510 min
Delta R.T.: -0.001 min
Response: 280972096
Conc: 747.39 ng/ml




#14 DINOSEB

R.T.: 11.150 min
Delta R.T.: -0.002 min
Response: 3896284686
Conc: 710.71 ng/ml



#14 DINOSEB


R.T.: 9.837 min
Delta R.T.: -0.002 min
Response: 1554607202
Conc: 761.38 ng/ml



#15 Picloram


R.T.: 10.968 min
Delta R.T.: -0.002 min
Response: 7287612188
Conc: 662.50 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 10.815 min
Delta R.T.: 0.000 min
Response: 2906268062
Conc: 698.52 ng/ml



#16 DCPA

R.T.: 11.417 min
Delta R.T.: -0.003 min
Response: 7680281964
Conc: 712.01 ng/ml



#16 DCPA

R.T.: 10.768 min
Delta R.T.: -0.002 min
Response: 3262066594
Conc: 758.55 ng/ml