

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072023\
 Data File : PS023773.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2023 15:20
 Operator : AR\AJ
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 19:00:46 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
 Quant Title : 8080.M
 QLast Update : Thu Jul 20 18:59:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.372	6.808	485.5E6	161.7E6	254.312	245.383
Target Compounds							
1) T	Dalapon	2.491	2.562	454.2E6	232.6E6	218.380	216.241
2) T	3,5-DICHL...	5.745	5.987	700.9E6	229.4E6	234.974	218.575
3) T	4-Nitroph...	6.210	6.419	267.4E6	89548264	223.204	217.504
5) T	DICAMBA	6.515	6.971	2014.3E6	642.8E6	227.783	213.329
8) T	DICHLORPROP	7.105	7.589	539.2E6	167.0E6	240.296	226.657
9) T	2,4-D	7.302	7.875	593.3E6	184.8E6	239.217	222.198
10) T	Pentachlo...	7.532	8.302	7761.6E6	2484.8E6	245.067	227.849
11) T	2,4,5-TP ...	8.055	8.661	2988.6E6	907.5E6	237.453	221.695
12) T	2,4,5-T	8.311	9.032	2652.7E6	843.8E6	230.679	217.365
13) T	2,4-DB	8.817	9.537	333.2E6	104.7E6	226.087	215.123
14) T	DINOSEB	9.873	9.865	1828.8E6	588.4E6	228.834	217.546
15) T	Picloram	9.697	10.823	3217.9E6	1236.6E6	212.601	204.329
16) T	DCPA	10.138	10.790	3460.9E6	1079.3E6	244.445	232.400

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072023\
Data File : PS023773.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2023 15:20
Operator : AR\AJ
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 19:00:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
Quant Title : 8080.M
QLast Update : Thu Jul 20 18:59:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

