

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072023\
 Data File : PS023775.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2023 16:09
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 20 19:01:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
 Quant Title : 8080.M
 QLast Update : Thu Jul 20 18:59:51 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.373	6.808	1431.7E6	494.3E6	750.000	750.000
Target Compounds							
1) T	Dalapon	2.490	2.561	1419.6E6	734.2E6	682.500	682.500
2) T	3,5-DICHL...	5.745	5.987	2080.6E6	731.9E6	697.500	697.500
3) T	4-Nitroph...	6.209	6.419	817.6E6	281.0E6	682.500	682.500
5) T	DICAMBA	6.515	6.972	6234.4E6	2124.5E6	705.000	705.000
8) T	DICHLORPROP	7.105	7.590	1582.1E6	519.4E6	705.000	705.000
9) T	2,4-D	7.302	7.875	1748.4E6	586.3E6	705.000	705.000
10) T	Pentachlo...	7.533	8.303	22565.8E6	7770.3E6	712.500	712.500
11) T	2,4,5-TP ...	8.055	8.662	8967.6E6	2916.6E6	712.500	712.500
12) T	2,4,5-T	8.310	9.033	8193.3E6	2765.9E6	712.500	712.500
13) T	2,4-DB	8.818	9.537	1049.9E6	346.7E6	712.500	712.500
14) T	DINOSEB	9.874	9.866	5634.3E6	1906.7E6	705.000	705.000
15) T	Picloram	9.697	10.824	10784.3E6	4312.0E6	712.500	712.500
16) T	DCPA	10.139	10.791	10194.1E6	3343.7E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072023\
Data File : PS023775.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2023 16:09
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 20 19:01:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
Quant Title : 8080.M
QLast Update : Thu Jul 20 18:59:51 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

