

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072122\
 Data File : PS020042.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jul 2022 18:44
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 22 01:15:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072022.M
 Quant Title : 8080.M
 QLast Update : Wed Jul 20 17:52:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.190	6.776	2672.6E6	696.9E6	806.013	773.525
Target Compounds							
1) T	Dalapon	3.011	2.524	3715.5E6	806.8E6	673.251	675.153
2) T	3,5-DICHL...	6.406	5.958	3216.0E6	992.3E6	740.178	715.604
3) T	4-Nitroph...	7.022	6.395	1137.6E6	342.0E6	791.365	707.833
5) T	DICAMBA	7.379	6.940	9923.9E6	2904.0E6	756.505	708.428
6) T	MCPD	7.549	7.037	823.1E6	226.0E6	78.511	72.338
7) T	MCPA	7.698	7.249	1146.5E6	312.2E6	76.751	73.195
8) T	DICHLORPROP	8.069	7.559	2622.9E6	752.4E6	757.189	725.223
9) T	2,4-D	8.300	7.851	3013.3E6	903.3E6	738.150	723.830
10) T	Pentachlo...	8.605	8.276	34740.0E6	9940.3E6	776.194	723.954
11) T	2,4,5-TP ...	9.155	8.637	15063.0E6	4341.4E6	749.135	731.293
12) T	2,4,5-T	9.440	9.015	15766.8E6	4097.4E6	729.558	742.976
13) T	2,4-DB	9.992	9.519	3168.8E6	541.8E6	727.489	758.959
14) T	DINOSEB	11.143	9.843	12350.1E6	2993.1E6	687.211	720.436
15) T	Picloram	10.970	10.824	24897.0E6	5565.1E6	714.747	724.172
16) T	DCPA	11.412	10.772	19595.5E6	5290.7E6	718.728	732.420

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072122\
Data File : PS020042.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jul 2022 18:44
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 22 01:15:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072022.M
Quant Title : 8080.M
QLast Update : Wed Jul 20 17:52:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

