

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072225\
 Data File : PS031173.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Jul 2025 10:11
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 01:53:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072125.M
 Quant Title : 8080.M
 QLast Update : Tue Jul 22 03:18:42 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.323	7.764	3209.7E6	731.1E6	738.152	720.359
Target Compounds							
1) T	Dalapon	2.689	2.703	4102.2E6	1826.1E6	653.955	643.743
2) T	3,5-DICHL...	6.485	6.711	3774.7E6	1022.0E6	683.477	663.682
3) T	4-Nitroph...	7.122	7.298	1169.1E6	1194.2E6	709.041	660.053
5) T	DICAMBA	7.512	7.966	11500.0E6	4457.3E6	697.096	690.699
6) T	MCPD	7.693	8.065	725.9E6	144.7E6	72.518	69.624
7) T	MCPA	7.843	8.314	892.8E6	215.0E6	71.345	68.154
8) T	DICHLORPROP	8.223	8.685	2607.8E6	1023.0E6	682.328	675.289
9) T	2,4-D	8.455	9.022	2634.9E6	1149.9E6	705.485	677.084
10) T	Pentachlo...	8.762	9.545	41070.3E6	27957.8E6	751.906	715.376
11) T	2,4,5-TP ...	9.339	9.926	15621.5E6	10389.3E6	711.571	697.539
12) T	2,4,5-T	9.634	10.353	14349.1E6	9904.0E6	734.804	696.507
13) T	2,4-DB	10.210	10.921	2243.1E6	812.8E6	750.224	694.373
14) T	DINOSEB	11.425	11.305	10877.6E6	7741.1E6	698.647	684.931
15) T	Picloram	11.235	12.416	15271.7E6	17885.1E6	763.336	718.420
16) T	DCPA	11.721	12.349	20621.6E6	16379.6E6	718.686	711.042

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072225\
Data File : PS031173.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Jul 2025 10:11
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 01:53:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072125.M
Quant Title : 8080.M
QLast Update : Tue Jul 22 03:18:42 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

