

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072420\
 Data File : PS011377.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2020 15:07
 Operator : DD\AJ
 Sample : L3379-02MSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 ETGI-361MSD

Manual Integrations
 APPROVED

Ankita
 7/27/2020 11:19:14 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 18:11:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.225	6.917	410.8E6	261.4E6	423.369	464.227
Target Compounds							
1) T	Dalapon	3.035	2.623	421.1E6	408.0E6	349.958	531.420m#
2) T	3,5-DICHL...	6.438	6.068	567.9E6	342.8E6	424.354	408.131
5) T	DICAMBA	7.416	7.085	1694.5E6	1012.3E6	437.185	442.038
6) T	MCPD	7.583	7.180	187.2E6	79444144	61.822	41.635 #
7) T	MCPA	7.731	7.395	179.8E6	126.3E6	42.166m	43.819m
8) T	DICHLORPROP	8.107	7.711	414.7E6	280.8E6	442.338	466.556
9) T	2,4-D	8.336	8.005	487.4E6	388.0E6	512.421	517.610
10) T	Pentachlo...	8.648	8.435	4089.5E6	2396.8E6	309.371	343.897
11) T	2,4,5-TP ...	9.196	8.792	2543.0E6	2634.4E6	480.198	861.675 #
12) T	2,4,5-T	9.482	9.175	2354.3E6	1502.6E6	466.588	495.417
13) T	2,4-DB	10.035	9.681	256.9E6	199.4E6	398.630	466.442
14) T	DINOSEB	11.195	10.014	883.4E6	524.0E6	229.173	244.314
15) T	Picloram	11.016	10.997	2242.7E6	1573.6E6	355.420	359.818
16) T	DCPA	11.461	10.942	2946.8E6	1889.9E6	457.366	499.708

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072420\
Data File : PS011377.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2020 15:07
Operator : DD\AJ
Sample : L3379-02MSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
ETGI-361MSD

Manual Integrations
APPROVED

Ankita
7/27/2020 11:19:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 18:11:25 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
Quant Title : 8080.M
QLast Update : Mon Jul 13 16:10:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

