

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072423\
 Data File : PS023808.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2023 09:28
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 07/25/2023
 Supervised By : Ankita Jodhani 07/25/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 13:59:22 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
 Quant Title : 8080.M
 QLast Update : Thu Jul 20 19:21:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.371	6.807	1418.1E6	479.4E6	715.249	704.908
Target Compounds							
1) T	Dalapon	2.491	2.562	1400.1E6	712.5E6	653.689	643.326
2) T	3,5-DICHL...	5.745	5.986	2072.2E6	711.2E6	671.006	664.169
3) T	4-Nitroph...	6.209	6.418	844.8E6	279.1E6	679.562	659.082
5) T	DICAMBA	6.514	6.971	6233.2E6	2078.2E6	688.028	680.985
8) T	DICHLORPROP	7.104	7.588	1576.9E6	500.6E6	674.968	662.627
9) T	2,4-D	7.300	7.874	1733.5E6	563.2E6	671.947	660.742
10) T	Pentachlo...	7.532	8.301	22648.8E6	7639.3E6	704.358	690.650
11) T	2,4,5-TP ...	8.054	8.660	9002.6E6	2861.3E6	696.163	688.628
12) T	2,4,5-T	8.309	9.031	8331.2E6	2716.8E6	707.426	690.119
13) T	2,4-DB	8.816	9.535	1063.8E6	350.5E6	701.889	706.922
14) T	DINOSEB	9.871	9.864	6222.0E6	1985.7E6	752.992m	718.287
15) T	Picloram	9.694	10.822	10978.0E6	4141.1E6	716.300	682.761
16) T	DCPA	10.137	10.789	10144.3E6	3302.0E6	695.097	694.505

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072423\
Data File : PS023808.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2023 09:28
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 07/25/2023
Supervised By : Ankita Jodhani 07/25/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 13:59:22 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
Quant Title : 8080.M
QLast Update : Thu Jul 20 19:21:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

