

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072618\
 Data File : PS001249.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2018 12:18
 Operator : UA
 Sample : PB111444BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 00:42:43 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
 Quant Title : 8080.M
 QLast Update : Sat Jul 21 01:18:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.619	11.037	92372675	83877802	497.832	461.184
Target Compounds							
1) T	Dalapon	3.010	2.424	86916817	88137622	366.543	386.785
2) T	3,5-DICHL...	10.036	9.185	108.2E6	99833431	400.974	391.889
3) T	4-Nitroph...	11.213	10.183	34125454	43746973	296.515	429.148 #
5) T	DICAMBA	11.935	11.324	287.8E6	292.8E6	400.310	427.204
6) T	MCPD	12.272	11.547	10525304	7005747	20.941	16.145
7) T	MCPA	12.522	11.922	29457852	9221861	37.353	14.182 #
8) T	DICHLORPROP	13.127	12.449	80633674	82353767	441.732	464.890
9) T	2,4-D	13.486	12.921	103.3E6	106.1E6	420.757	447.147
10) T	Pentachlo...	13.874	13.476	1223.9E6	1178.3E6	434.687	428.138
11) T	2,4,5-TP ...	14.744	14.090	463.0E6	449.7E6	412.548	411.285
12) T	2,4,5-T	15.141	14.661	455.9E6	376.4E6	409.753	407.444
13) T	2,4-DB	15.893	15.307	61868535	56146824	429.751	421.642
14) T	DINOSEB	17.392	15.738	307.8E6	283.4E6	400.698	398.754
15) T	Picloram	17.151	16.921	522.6E6	414.5E6	330.655	300.703
16) T	DCPA	17.733	16.977	374.9E6	492.3E6	277.704	340.890



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072618\
Data File : PS001249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2018 12:18
Operator : UA
Sample : PB111444BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 00:42:43 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
Quant Title : 8080.M
QLast Update : Sat Jul 21 01:18:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.010 min
Delta R.T.: -0.003 min
Response: 86916817
Conc: 366.54 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.424 min
Delta R.T.: -0.002 min
Response: 88137622
Conc: 386.79 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.036 min
Delta R.T.: 0.000 min
Response: 108221252
Conc: 400.97 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.185 min
Delta R.T.: -0.001 min
Response: 99833431
Conc: 391.89 ng/ml



#3 4-Nitrophenol


R.T.: 11.213 min
Delta R.T.: 0.000 min
Response: 34125454
Conc: 296.51 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 10.183 min
Delta R.T.: -0.003 min
Response: 43746973
Conc: 429.15 ng/ml




#4 2,4-DCAA

R.T.: 11.619 min
Delta R.T.: 0.000 min
Response: 92372675
Conc: 497.83 ng/ml



#4 2,4-DCAA


R.T.: 11.037 min
Delta R.T.: -0.002 min
Response: 83877802
Conc: 461.18 ng/ml



#5 DICAMBA


R.T.: 11.935 min
Delta R.T.: 0.001 min
Response: 287812709
Conc: 400.31 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 11.324 min
Delta R.T.: -0.002 min
Response: 292797982
Conc: 427.20 ng/ml




#6 MCP

R.T.: 12.272 min
Delta R.T.: 0.002 min
Response: 10525304
Conc: 20.94 ug/ml



#6 MCP


R.T.: 11.547 min
Delta R.T.: 0.000 min
Response: 7005747
Conc: 16.15 ug/ml



#7 MCPA


R.T.: 12.522 min
Delta R.T.: 0.001 min
Response: 29457852
Conc: 37.35 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 11.922 min
Delta R.T.: 0.000 min
Response: 9221861
Conc: 14.18 ug/ml




#8 DICHLORPROP

R.T.: 13.127 min
Delta R.T.: 0.000 min
Response: 80633674
Conc: 441.73 ng/ml



#8 DICHLORPROP


R.T.: 12.449 min
Delta R.T.: -0.002 min
Response: 82353767
Conc: 464.89 ng/ml



#9 2,4-D


R.T.: 13.486 min
Delta R.T.: 0.000 min
Response: 103281844
Conc: 420.76 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 12.921 min
Delta R.T.: -0.002 min
Response: 106104789
Conc: 447.15 ng/ml




#10 Pentachlorophenol

R.T.: 13.874 min
Delta R.T.: 0.000 min
Response: 1223865538
Conc: 434.69 ng/ml



#10 Pentachlorophenol


R.T.: 13.476 min
Delta R.T.: -0.001 min
Response: 1178311781
Conc: 428.14 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.744 min
Delta R.T.: 0.000 min
Response: 463011455
Conc: 412.55 ng/ml

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 14.090 min
Delta R.T.: -0.001 min
Response: 449720291
Conc: 411.28 ng/ml




#12 2,4,5-T

R.T.: 15.141 min
Delta R.T.: 0.000 min
Response: 455946460
Conc: 409.75 ng/ml



#12 2,4,5-T


R.T.: 14.661 min
Delta R.T.: 0.000 min
Response: 376367804
Conc: 407.44 ng/ml



#13 2,4-DB


R.T.: 15.893 min
Delta R.T.: 0.000 min
Response: 61868535
Conc: 429.75 ng/ml

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 15.307 min
Delta R.T.: -0.001 min
Response: 56146824
Conc: 421.64 ng/ml




#14 DINOSEB

R.T.: 17.392 min
Delta R.T.: 0.000 min
Response: 307762094
Conc: 400.70 ng/ml



#14 DINOSEB


R.T.: 15.738 min
Delta R.T.: 0.000 min
Response: 283437606
Conc: 398.75 ng/ml



#15 Picloram


R.T.: 17.151 min
Delta R.T.: 0.000 min
Response: 522648054
Conc: 330.65 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 16.921 min
Delta R.T.: -0.001 min
Response: 414521945
Conc: 300.70 ng/ml



#16 DCPA

R.T.: 17.733 min
Delta R.T.: 0.001 min
Response: 374923118
Conc: 277.70 ng/ml



#16 DCPA

R.T.: 16.977 min
Delta R.T.: -0.001 min
Response: 492345011
Conc: 340.89 ng/ml