

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072618\
 Data File : PS001268.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2018 22:07
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 7/27/2018 3:35:56 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 00:50:49 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
 Quant Title : 8080.M
 QLast Update : Sat Jul 21 01:18:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.657	11.076	135.6E6	129.5E6	730.674	712.008m
Target Compounds							
1) T	Dalapon	3.051	2.464	154.6E6	150.1E6	652.035	658.810m
2) T	3,5-DICHL...	10.073	9.224	182.9E6	171.4E6	677.587	672.948
3) T	4-Nitroph...	11.250	10.222	75020234	67018214	651.848	657.434
5) T	DICAMBA	11.972	11.364	502.1E6	475.0E6	698.390	693.109
6) T	MCPD	12.310	11.588	35286911	31458391	70.208m	72.499m
7) T	MCPA	12.561	11.964	56830503	45888682	72.062	70.571m
8) T	DICHLORPROP	13.165	12.488	125.0E6	119.4E6	684.783m	674.109
9) T	2,4-D	13.522	12.960	168.9E6	158.9E6	688.227	669.732m
10) T	Pentachlo...	13.912	13.515	2004.4E6	1956.1E6	711.925	710.749m
11) T	2,4,5-TP ...	14.781	14.128	792.7E6	768.8E6	706.346	703.068
12) T	2,4,5-T	15.178	14.660	782.7E6	753.2E6	703.373	815.417
13) T	2,4-DB	15.929	15.345	101.1E6	93403958	701.951	701.429
14) T	DINOSEB	17.428	15.777	544.4E6	500.1E6	708.819	703.514
15) T	Picloram	17.186	16.959	1114.8E6	939.5E6	705.308	681.523
16) T	DCPA	17.768	16.959	970.8E6	939.5E6	719.095	650.483

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072618\
Data File : PS001268.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2018 22:07
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
7/27/2018 3:35:56 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 00:50:49 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
Quant Title : 8080.M
QLast Update : Sat Jul 21 01:18:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

