

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072621\
 Data File : PS016182.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2021 09:40
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 02:36:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 30 00:31:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.186	6.765	1110.3E6	384.9E6	723.203	749.831
Target Compounds							
1) T	Dalapon	3.015	2.517	1488.6E6	485.0E6	686.349	693.873
2) T	3,5-DICHL...	6.404	5.957	1531.5E6	559.3E6	706.117	726.562
3) T	4-Nitroph...	7.014	6.386	562.7E6	213.3E6	636.337	675.711
5) T	DICAMBA	7.376	6.928	3947.6E6	1516.7E6	689.040	720.045
6) T	MCPD	7.546	7.025	310.0E6	121.3E6	70.704	73.095
7) T	MCPA	7.695	7.237	444.1E6	183.7E6	69.843	69.695
8) T	DICHLORPROP	8.069	7.548	939.1E6	371.3E6	675.151	705.184
9) T	2,4-D	8.297	7.839	1057.3E6	439.9E6	659.554	697.644
10) T	Pentachlo...	8.606	8.266	14458.3E6	5366.1E6	729.595	745.991
11) T	2,4,5-TP ...	9.157	8.627	5080.1E6	2052.3E6	684.620	727.353
12) T	2,4,5-T	9.440	9.003	5243.2E6	2065.7E6	671.460	716.079
13) T	2,4-DB	9.990	9.509	1039.8E6	266.3E6	625.990	708.495
14) T	DINOSEB	11.148	9.836	4146.8E6	1577.9E6	756.407	772.801
15) T	Picloram	10.968	10.814	7418.7E6	2994.0E6	674.412	719.620
16) T	DCPA	11.415	10.766	7668.0E6	3214.3E6	710.871	747.445

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072621\
Data File : PS016182.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2021 09:40
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 02:36:53 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
Quant Title : 8080.M
QLast Update : Wed Jun 30 00:31:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

