

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072720\
 Data File : PS011397.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jul 2020 09:43
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 27 14:21:37 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.225	6.918	755.6E6	434.5E6	778.704	771.758
Target Compounds							
1) T	Dalapon	3.037	2.623	814.2E6	494.2E6	676.634	643.783
2) T	3,5-DICHL...	6.438	6.068	973.0E6	604.8E6	727.153	720.102
3) T	4-Nitroph...	7.055	6.524	309.3E6	228.0E6	701.571	695.285
5) T	DICAMBA	7.416	7.086	2872.2E6	1679.3E6	741.042	733.334
6) T	MCPD	7.586	7.184	221.3E6	140.9E6	73.089	73.846
7) T	MCPA	7.735	7.400	311.1E6	209.4E6	72.952	72.663
8) T	DICHLORPROP	8.108	7.712	675.3E6	436.8E6	720.269	725.585
9) T	2,4-D	8.337	8.006	676.2E6	538.9E6	710.997	718.873
10) T	Pentachlo...	8.648	8.436	9687.3E6	5014.3E6	732.849	719.441
11) T	2,4,5-TP ...	9.197	8.796	3864.2E6	2267.4E6	729.670	741.640
12) T	2,4,5-T	9.483	9.176	3624.5E6	2214.4E6	718.304	730.087
13) T	2,4-DB	10.036	9.682	441.5E6	320.2E6	684.998	749.048
14) T	DINOSEB	11.196	10.014	2814.7E6	1628.9E6	730.202	759.400
15) T	Picloram	11.019	10.999	4205.3E6	2982.6E6	666.460	681.982
16) T	DCPA	11.461	10.943	4577.8E6	2816.6E6	710.518	744.749

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072720\
Data File : PS011397.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2020 09:43
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 27 14:21:37 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
Quant Title : 8080.M
QLast Update : Mon Jul 13 16:10:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

