

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072720\
 Data File : PS011409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Jul 2020 15:27
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 02:15:10 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.225	6.918	759.9E6	438.3E6	783.157	778.430
Target Compounds							
1) T	Dalapon	3.036	2.621	765.8E6	469.6E6	636.363	611.704
2) T	3,5-DICHL...	6.438	6.068	977.4E6	609.8E6	730.434	726.032
3) T	4-Nitroph...	7.056	6.524	315.4E6	231.7E6	715.533	706.555
5) T	DICAMBA	7.416	7.087	2887.6E6	1693.6E6	745.021	739.572
6) T	MCPD	7.586	7.184	224.0E6	142.2E6	73.968	74.542
7) T	MCPA	7.736	7.400	315.5E6	214.5E6	73.969	74.432
8) T	DICHLORPROP	8.108	7.712	677.9E6	439.3E6	723.037	729.869
9) T	2,4-D	8.338	8.007	682.0E6	555.4E6	717.055	740.876
10) T	Pentachlo...	8.649	8.437	9706.2E6	5190.0E6	734.278	744.655
11) T	2,4,5-TP ...	9.197	8.797	3875.5E6	2282.3E6	731.808	746.502
12) T	2,4,5-T	9.484	9.176	3630.7E6	2235.2E6	719.547	736.953
13) T	2,4-DB	10.036	9.683	449.3E6	324.9E6	697.040	759.897
14) T	DINOSEB	11.195	10.015	2814.8E6	1622.5E6	730.240	756.443
15) T	Picloram	11.019	11.000	4056.8E6	2930.3E6	642.928	670.026
16) T	DCPA	11.461	10.944	4609.1E6	2842.9E6	715.375	751.713

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072720\
Data File : PS011409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Jul 2020 15:27
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 02:15:10 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
Quant Title : 8080.M
QLast Update : Mon Jul 13 16:10:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

