

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072822\
 Data File : PS020101.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2022 15:54
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 22:10:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072022.M
 Quant Title : 8080.M
 QLast Update : Wed Jul 20 17:52:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.188	6.774	2463.5E6	666.5E6	742.943	739.858
Target Compounds							
1) T	Dalapon	3.013	2.526	3567.6E6	776.5E6	646.443	649.811
2) T	3,5-DICHL...	6.404	5.957	2922.1E6	954.7E6	672.539	688.475
3) T	4-Nitroph...	7.019	6.394	1118.8E6	344.4E6	778.247	712.895
5) T	DICAMBA	7.378	6.939	9106.6E6	2773.6E6	694.203	676.608
6) T	MCPD	7.547	7.036	734.9E6	212.8E6	70.100	68.125
7) T	MCPA	7.696	7.248	968.4E6	294.2E6	64.829	68.974
8) T	DICHLORPROP	8.068	7.558	2261.7E6	708.0E6	652.897	682.510
9) T	2,4-D	8.297	7.850	2639.6E6	854.5E6	646.592	684.712
10) T	Pentachlo...	8.604	8.275	30556.1E6	9536.2E6	682.712	694.524
11) T	2,4,5-TP ...	9.153	8.636	13215.3E6	4094.3E6	657.241	689.658
12) T	2,4,5-T	9.438	9.013	13755.4E6	3887.2E6	636.488	704.867
13) T	2,4-DB	9.989	9.517	2968.0E6	512.3E6	681.388	717.696
14) T	DINOSEB	11.142	9.842	11780.1E6	2867.7E6	655.491	690.259
15) T	Picloram	10.967	10.822	22964.5E6	4944.3E6	659.270	643.384
16) T	DCPA	11.411	10.770	18667.2E6	4974.6E6	684.677	688.666

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072822\
Data File : PS020101.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2022 15:54
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 22:10:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072022.M
Quant Title : 8080.M
QLast Update : Wed Jul 20 17:52:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

