

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072822\
 Data File : PS020113.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2022 20:46
 Operator : AJ\MA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 28 22:13:52 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072022.M
 Quant Title : 8080.M
 QLast Update : Wed Jul 20 17:52:40 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.189	6.775	2495.6E6	675.3E6	752.636	749.627
Target Compounds							
1) T	Dalapon	3.013	2.527	3569.1E6	780.0E6	646.716	652.691
2) T	3,5-DICHL...	6.405	5.957	2951.4E6	961.2E6	679.269	693.195
3) T	4-Nitroph...	7.020	6.394	1156.4E6	351.5E6	804.406	727.615
5) T	DICAMBA	7.378	6.939	9215.9E6	2799.3E6	702.534	682.888
6) T	MCPD	7.548	7.036	743.4E6	214.5E6	70.909	68.660
7) T	MCPA	7.696	7.248	1010.0E6	294.1E6	67.609	68.960
8) T	DICHLORPROP	8.068	7.558	2313.4E6	716.7E6	667.825	690.884
9) T	2,4-D	8.297	7.850	2698.4E6	864.7E6	660.992	692.901
10) T	Pentachlo...	8.604	8.276	31241.1E6	9648.7E6	698.017	702.719
11) T	2,4,5-TP ...	9.153	8.636	13464.5E6	4134.7E6	669.633	696.479
12) T	2,4,5-T	9.438	9.013	14067.0E6	3933.0E6	650.905	713.163
13) T	2,4-DB	9.989	9.517	2988.0E6	518.7E6	685.980	726.607
14) T	DINOSEB	11.142	9.842	11881.2E6	2908.2E6	661.115	700.001
15) T	Picloram	10.967	10.822	23138.1E6	4985.5E6	664.253	648.745
16) T	DCPA	11.410	10.771	18781.7E6	5031.9E6	688.876	696.592

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS072822\
Data File : PS020113.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2022 20:46
Operator : AJ\MA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 28 22:13:52 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072022.M
Quant Title : 8080.M
QLast Update : Wed Jul 20 17:52:40 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

