

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073018\
 Data File : PS001312.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jul 2018 06:36
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 7/31/2018 1:45:04 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 31 07:04:02 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
 Quant Title : 8080.M
 QLast Update : Sat Jul 21 01:18:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.634	11.054	140.9E6	138.9E6	759.337	763.867m
Target Compounds							
1) T	Dalapon	3.029	2.446	152.1E6	151.7E6	641.466m	665.687
2) T	3,5-DICHL...	10.051	9.203	203.8E6	192.5E6	755.109	755.741
3) T	4-Nitroph...	11.227	10.201	127.3E6	121.5E6	1106.225	1191.610
5) T	DICAMBA	11.950	11.342	501.1E6	481.9E6	696.930	703.078
6) T	MCPD	12.288	11.566	50891928	44654365	101.256m	102.911m
7) T	MCPA	12.539	11.942	74408859	59255836	94.352	91.128m
8) T	DICHLORPROP	13.141	12.467	130.6E6	124.7E6	715.277	704.216
9) T	2,4-D	13.500	12.939	172.5E6	167.8E6	702.680	707.043
10) T	Pentachlo...	13.889	13.493	2004.4E6	1985.0E6	711.922	721.247
11) T	2,4,5-TP ...	14.757	14.107	789.2E6	773.0E6	703.220	706.907
12) T	2,4,5-T	15.155	14.677	788.6E6	626.3E6	708.676	678.030
13) T	2,4-DB	15.906	15.324	103.7E6	93911506	720.019	705.240
14) T	DINOSEB	17.406	15.755	547.2E6	510.2E6	712.447	717.778
15) T	Picloram	17.163	16.938	1082.5E6	965.8E6	684.867m	700.588
16) T	DCPA	17.745	16.994	980.4E6	1007.9E6	726.149	697.858

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073018\
Data File : PS001312.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jul 2018 06:36
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
7/31/2018 1:45:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 07:04:02 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
Quant Title : 8080.M
QLast Update : Sat Jul 21 01:18:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

