

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073020\
 Data File : PS011524.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2020 20:05
 Operator : DD\AJ
 Sample : L3183-22MS
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 OK-02-072820MS

Manual Integrations
 APPROVED

Ankita
 7/31/2020 9:42:09 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 31 02:16:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.224	6.917	551.1E6	347.9E6	567.969m	617.838
Target Compounds							
1) T	Dalapon	3.037	2.625	486.5E6	353.6E6	404.267	460.641
2) T	3,5-DICHL...	6.438	6.068	662.4E6	389.4E6	495.016	463.646
5) T	DICAMBA	7.415	7.084	2041.5E6	1173.6E6	526.733	512.485m
6) T	MCP	7.583	7.181	218.5E6	95112199	72.147m	49.847 #
7) T	MCPA	7.732	7.395	275.6E6	203.4E6	64.627	70.553
8) T	DICHLORPROP	8.107	7.710	492.2E6	327.2E6	524.961	543.505
9) T	2,4-D	8.336	8.005	581.2E6	453.3E6	611.052	604.662
10) T	Pentachlo...	8.647	8.434	4731.9E6	2730.9E6	357.968	391.830
11) T	2,4,5-TP ...	9.196	8.792	3033.1E6	2962.4E6	572.739	968.977 #
12) T	2,4,5-T	9.482	9.174	2801.2E6	1759.2E6	555.150	580.018
13) T	2,4-DB	10.035	9.681	302.1E6	236.5E6	468.728	553.292
14) T	DINOSEB	11.195	10.014	1049.1E6	598.7E6	272.156	279.105
15) T	Picloram	11.017	10.998	2155.0E6	1479.2E6	341.533	338.239
16) T	DCPA	11.461	10.942	3224.2E6	2103.4E6	500.421	556.172

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073020\
Data File : PS011524.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2020 20:05
Operator : DD\AJ
Sample : L3183-22MS
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
OK-02-072820MS

Manual Integrations
APPROVED

Ankita
7/31/2020 9:42:09 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 02:16:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
Quant Title : 8080.M
QLast Update : Mon Jul 13 16:10:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

