

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073021\
 Data File : PS016240.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2021 15:19
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 15:49:58 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 30 00:31:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.184	6.765	1151.0E6	404.2E6	749.725	787.388
Target Compounds							
1) T	Dalapon	3.015	2.517	1502.3E6	496.1E6	692.651	709.829
2) T	3,5-DICHL...	6.403	5.956	1538.0E6	579.5E6	709.085	752.897
3) T	4-Nitroph...	7.013	6.386	584.1E6	224.4E6	660.553	711.116
5) T	DICAMBA	7.375	6.928	4133.1E6	1596.0E6	721.410	757.692
6) T	MCPD	7.545	7.025	315.1E6	127.2E6	71.883	76.684
7) T	MCPA	7.694	7.237	450.7E6	193.8E6	70.880	73.540
8) T	DICHLORPROP	8.068	7.548	976.2E6	392.9E6	701.821	746.155
9) T	2,4-D	8.297	7.839	1111.8E6	465.0E6	693.563	737.413
10) T	Pentachlo...	8.605	8.266	14290.3E6	5427.8E6	721.117	754.562
11) T	2,4,5-TP ...	9.155	8.627	5292.8E6	2147.9E6	713.283	761.251
12) T	2,4,5-T	9.439	9.004	5548.2E6	2162.9E6	710.511	749.768
13) T	2,4-DB	9.989	9.509	1125.6E6	283.7E6	677.618	754.725
14) T	DINOSEB	11.146	9.837	3961.1E6	1578.5E6	722.535	773.083
15) T	Picloram	10.967	10.815	7037.1E6	2842.5E6	639.722	683.191
16) T	DCPA	11.415	10.767	7727.3E6	3310.4E6	716.363	769.789

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073021\
Data File : PS016240.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2021 15:19
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 15:49:58 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
Quant Title : 8080.M
QLast Update : Wed Jun 30 00:31:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

