

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073021\
 Data File : PS016247.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2021 18:08
 Operator : DD\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 31 01:55:36 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
 Quant Title : 8080.M
 QLast Update : Wed Jun 30 00:31:48 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.185	6.765	1170.6E6	408.4E6	762.506	795.688
Target Compounds							
1) T	Dalapon	3.017	2.520	1511.5E6	499.2E6	696.926	714.241
2) T	3,5-DICHL...	6.403	5.957	1546.1E6	583.1E6	712.853	757.571
3) T	4-Nitroph...	7.013	6.386	591.5E6	227.6E6	668.832	721.072
5) T	DICAMBA	7.376	6.928	4218.7E6	1616.5E6	736.354	767.405
6) T	MCPD	7.545	7.026	323.1E6	129.5E6	73.698	78.027
7) T	MCPA	7.694	7.238	463.5E6	196.1E6	72.887	74.412
8) T	DICHLORPROP	8.068	7.549	1000.3E6	398.8E6	719.157	757.258
9) T	2,4-D	8.297	7.839	1134.4E6	472.4E6	707.648	749.096
10) T	Pentachlo...	8.605	8.266	14507.6E6	5491.4E6	732.085	763.404
11) T	2,4,5-TP ...	9.155	8.627	5384.5E6	2179.5E6	725.635	772.460
12) T	2,4,5-T	9.439	9.004	5638.4E6	2204.0E6	722.063	764.028
13) T	2,4-DB	9.989	9.510	1146.6E6	289.3E6	690.279	769.551
14) T	DINOSEB	11.147	9.836	3977.6E6	1606.7E6	725.550	786.913
15) T	Picloram	10.966	10.815	7058.7E6	2878.6E6	641.688	691.875
16) T	DCPA	11.415	10.767	7793.1E6	3369.7E6	722.463	783.584

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073021\
Data File : PS016247.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2021 18:08
Operator : DD\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 01:55:36 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS063021.M
Quant Title : 8080.M
QLast Update : Wed Jun 30 00:31:48 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

