

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073120\
 Data File : PS011547.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 31 Jul 2020 14:52
 Operator : DD\AJ
 Sample : L3435-05MSD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SP-3MSD

Manual Integrations
 APPROVED

Ankita
 8/3/2020 12:08:53 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 31 18:45:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.225	6.917	442.3E6	258.4E6	455.778	458.967m
Target Compounds							
1) T	Dalapon	3.032	2.619	224.6E6	188.8E6	186.660	245.992m#
2) T	3,5-DICHL...	6.438	6.069	343.3E6	200.2E6	256.530	238.334
3) T	4-Nitroph...	7.061	6.530	9653762	4449930	21.898m	13.572m#
5) T	DICAMBA	7.416	7.085	952.8E6	565.4E6	245.838	246.921
6) T	MCPD	7.581	7.177	78543795	44793727	25.938	23.476
7) T	MCPA	7.729	7.391	112.3E6	72292773	26.324	25.081m
8) T	DICHLORPROP	8.108	7.711	239.3E6	158.9E6	255.200	264.024
9) T	2,4-D	8.338	8.006	263.6E6	213.5E6	277.108	284.790
10) T	Pentachlo...	8.648	8.435	3601.3E6	2072.3E6	272.441	297.326
11) T	2,4,5-TP ...	9.197	8.795	1393.5E6	862.3E6	263.127	282.045
12) T	2,4,5-T	9.482	9.176	1267.6E6	825.9E6	251.218	272.288
13) T	2,4-DB	10.036	9.681	152.0E6	126.1E6	235.867	294.956m#
14) T	DINOSEB	11.195	10.016	99155236	56131778	25.723	26.170
15) T	Picloram	11.019	10.999	1398.9E6	1001.9E6	221.697	229.102
16) T	DCPA	11.462	10.942	1583.6E6	1161.6E6	245.792	307.144

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS073120\
Data File : PS011547.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 31 Jul 2020 14:52
Operator : DD\AJ
Sample : L3435-05MSD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampled :
SP-3MSD

Manual Integrations
APPROVED

Ankita
8/3/2020 12:08:53 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 31 18:45:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
Quant Title : 8080.M
QLast Update : Mon Jul 13 16:10:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

