

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080118\
 Data File : PS001326.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Aug 2018 17:30
 Operator : UA
 Sample : PB111622BS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 00:47:21 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
 Quant Title : 8080.M
 QLast Update : Sat Jul 21 01:18:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.618	11.037	93883758	87957521	505.976	483.616
Target Compounds							
1) T	Dalapon	3.009	2.425	85752462	84060014	361.633	368.891
2) T	3,5-DICHL...	10.033	9.183	113.4E6	114.3E6	420.028	448.667
3) T	4-Nitroph...	11.210	10.183	41680029	53193064	362.156	521.812 #
5) T	DICAMBA	11.933	11.324	305.4E6	313.9E6	424.829	458.011
7) T	MCPA	12.521	0.000	30230044	0	38.332	N.D. #
8) T	DICHLORPROP	13.126	12.449	85100103	83235425	466.200	469.867
9) T	2,4-D	13.483	12.920	111.7E6	111.2E6	455.205	468.602
10) T	Pentachlo...	13.873	13.475	1325.0E6	1260.7E6	470.605	458.091
11) T	2,4,5-TP ...	14.742	14.089	501.5E6	481.3E6	446.825	440.206
12) T	2,4,5-T	15.140	14.659	497.4E6	412.0E6	447.024	446.057
13) T	2,4-DB	15.891	15.305	66571680	60272813	462.420	452.626
14) T	DINOSEB	17.390	15.737	336.6E6	313.5E6	438.261	441.083
15) T	Picloram	17.149	16.919	593.5E6	441.0E6	375.453	319.931
16) T	DCPA	17.731	16.976	417.2E6	548.4E6	309.049	379.669



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080118\
Data File : PS001326.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 Aug 2018 17:30
Operator : UA
Sample : PB111622BS
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 00:47:21 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
Quant Title : 8080.M
QLast Update : Sat Jul 21 01:18:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Dalapon


R.T.: 3.009 min
Delta R.T.: -0.004 min
Response: 85752462
Conc: 361.63 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Dalapon

R.T.: 2.425 min
Delta R.T.: 0.000 min
Response: 84060014
Conc: 368.89 ng/ml




#2 3,5-DICHLOROBENZOIC ACID

R.T.: 10.033 min
Delta R.T.: -0.002 min
Response: 113363886
Conc: 420.03 ng/ml



#2 3,5-DICHLOROBENZOIC ACID


R.T.: 9.183 min
Delta R.T.: -0.003 min
Response: 114297412
Conc: 448.67 ng/ml



#3 4-Nitrophenol


R.T.: 11.210 min
Delta R.T.: -0.003 min
Response: 41680029
Conc: 362.16 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 10.183 min
Delta R.T.: -0.003 min
Response: 53193064
Conc: 521.81 ng/ml




#4 2,4-DCAA

R.T.: 11.618 min
Delta R.T.: -0.002 min
Response: 93883758
Conc: 505.98 ng/ml



#4 2,4-DCAA


R.T.: 11.037 min
Delta R.T.: -0.002 min
Response: 87957521
Conc: 483.62 ng/ml



#5 DICAMBA


R.T.: 11.933 min
Delta R.T.: 0.000 min
Response: 305441128
Conc: 424.83 ng/ml

Instrument :
ECD_S
ClientSampleId :




#5 DICAMBA

R.T.: 11.324 min
Delta R.T.: -0.002 min
Response: 313913054
Conc: 458.01 ng/ml




#6 MCP

R.T.: 12.270 min
Delta R.T.: 0.000 min
Response: 10427009
Conc: 20.75 ug/ml



#6 MCP


R.T.: 11.596 min
Delta R.T.: 0.050 min
Response: -9230801
Conc: N.D.



#7 MCPA


R.T.: 12.521 min
Delta R.T.: 0.000 min
Response: 30230044
Conc: 38.33 ug/ml

Instrument :
ECD_S
ClientSampleId :




#7 MCPA

R.T.: 0.000 min
Exp R.T. : 11.922 min
Response: 0
Conc: N.D.




#8 DICHLORPROP

R.T.: 13.126 min
Delta R.T.: -0.002 min
Response: 85100103
Conc: 466.20 ng/ml



#8 DICHLORPROP


R.T.: 12.449 min
Delta R.T.: -0.003 min
Response: 83235425
Conc: 469.87 ng/ml



#9 2,4-D


R.T.: 13.483 min
Delta R.T.: -0.003 min
Response: 111737679
Conc: 455.20 ng/ml

Instrument :
ECD_S
ClientSampleId :




#9 2,4-D

R.T.: 12.920 min
Delta R.T.: -0.002 min
Response: 111195963
Conc: 468.60 ng/ml




#10 Pentachlorophenol

R.T.: 13.873 min
Delta R.T.: -0.002 min
Response: 1324992656
Conc: 470.61 ng/ml



#10 Pentachlorophenol


R.T.: 13.475 min
Delta R.T.: -0.003 min
Response: 1260749649
Conc: 458.09 ng/ml



#11 2,4,5-TP (SILVEX)


R.T.: 14.742 min
Delta R.T.: -0.001 min
Response: 501480822
Conc: 446.82 ng/ml

Instrument :
ECD_S
ClientSampleId :




#11 2,4,5-TP (SILVEX)

R.T.: 14.089 min
Delta R.T.: -0.002 min
Response: 481344844
Conc: 440.21 ng/ml




#12 2,4,5-T

R.T.: 15.140 min
Delta R.T.: -0.002 min
Response: 497419667
Conc: 447.02 ng/ml



#12 2,4,5-T


R.T.: 14.659 min
Delta R.T.: -0.002 min
Response: 412035225
Conc: 446.06 ng/ml



#13 2,4-DB


R.T.: 15.891 min
Delta R.T.: -0.002 min
Response: 66571680
Conc: 462.42 ng/ml

Instrument :
ECD_S
ClientSampleId :




#13 2,4-DB

R.T.: 15.305 min
Delta R.T.: -0.003 min
Response: 60272813
Conc: 452.63 ng/ml




#14 DINOSEB

R.T.: 17.390 min
Delta R.T.: -0.002 min
Response: 336612628
Conc: 438.26 ng/ml



#14 DINOSEB


R.T.: 15.737 min
Delta R.T.: -0.003 min
Response: 313525385
Conc: 441.08 ng/ml



#15 Picloram


R.T.: 17.149 min
Delta R.T.: -0.003 min
Response: 593458384
Conc: 375.45 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 16.919 min
Delta R.T.: -0.003 min
Response: 441027993
Conc: 319.93 ng/ml



#16 DCPA

R.T.: 17.731 min
Delta R.T.: 0.000 min
Response: 417241374
Conc: 309.05 ng/ml



#16 DCPA

R.T.: 16.976 min
Delta R.T.: -0.002 min
Response: 548352093
Conc: 379.67 ng/ml