

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080118\
 Data File : PS001344.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2018 02:48
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Sohil
 8/2/2018 5:55:27 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 04:11:07 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
 Quant Title : 8080.M
 QLast Update : Sat Jul 21 01:18:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.657	11.077	146.8E6	139.4E6	790.947	766.480m
Target Compounds							
1) T	Dalapon	3.052	2.467	163.8E6	161.2E6	690.769m	707.462m
2) T	3,5-DICHL...	10.073	9.225	198.9E6	182.6E6	736.846	716.710
3) T	4-Nitroph...	11.249	10.223	83954693	73774468	729.479	723.712
5) T	DICAMBA	11.972	11.364	534.6E6	494.4E6	743.571	721.290
6) T	MCPD	12.311	11.588	40608950	34326377	80.797m	79.109m
7) T	MCPA	12.562	11.965	62566372	49462032	79.335	76.067
8) T	DICHLORPROP	13.164	12.489	135.0E6	126.4E6	739.363m	713.624
9) T	2,4-D	13.523	12.960	178.9E6	167.5E6	728.742m	705.807
10) T	Pentachlo...	13.912	13.516	2149.6E6	2065.6E6	763.470	750.550m
11) T	2,4,5-TP ...	14.781	14.129	844.1E6	802.2E6	752.078	733.660
12) T	2,4,5-T	15.178	14.698	839.4E6	624.1E6	754.337	675.593m
13) T	2,4-DB	15.929	15.345	109.0E6	97508053	756.828	732.249
14) T	DINOSEB	17.429	15.777	581.0E6	530.3E6	756.449	746.010
15) T	Picloram	17.187	16.959	1133.3E6	976.6E6	716.992	708.476
16) T	DCPA	17.768	16.959	1022.9E6	976.6E6	757.646	676.208

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS001344.D
Data File : PS001344.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2018 02:48
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Sohil
8/2/2018 5:55:27 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 04:11:07 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
Quant Title : 8080.M
QLast Update : Sat Jul 21 01:18:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

