

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080221\
 Data File : PS016262.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2021 13:37
 Operator : DD\AJ
 Sample : HSTDICC1500
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 14:01:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080221.M
 Quant Title : 8080.M
 QLast Update : Mon Aug 02 13:59:52 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.186	6.766	2164.2E6	785.2E6	1423.817	1451.498
Target Compounds							
1) T	Dalapon	3.015	2.517	2809.9E6	939.9E6	1335.187	1322.783
2) T	3,5-DICHL...	6.404	5.957	2826.9E6	1087.9E6	1302.473	1314.247
3) T	4-Nitroph...	7.013	6.386	1104.4E6	418.6E6	1330.127	1303.466
5) T	DICAMBA	7.377	6.930	7609.7E6	2965.2E6	1300.709	1299.892
6) T	MCPD	7.552	7.032	632.0E6	252.9E6	142.871	140.201
7) T	MCPA	7.702	7.246	877.8E6	370.7E6	137.303	134.226
8) T	DICHLORPROP	8.069	7.549	1851.3E6	760.0E6	1335.251	1357.966
9) T	2,4-D	8.297	7.839	2090.0E6	896.4E6	1339.534	1361.722
10) T	Pentachlo...	8.607	8.267	24673.9E6	9134.0E6	1211.637	1184.996
11) T	2,4,5-TP ...	9.157	8.629	9714.4E6	3956.3E6	1291.137	1297.724
12) T	2,4,5-T	9.440	9.005	10135.4E6	3984.5E6	1283.955	1306.594
13) T	2,4-DB	9.989	9.510	2196.6E6	556.2E6	1435.284	1407.052
14) T	DINOSEB	11.148	9.838	7134.2E6	2856.9E6	1305.222	1330.927
15) T	Picloram	10.967	10.815	14420.3E6	5749.7E6	1351.938	1325.487
16) T	DCPA	11.415	10.768	13907.8E6	5949.7E6	1274.259	1271.635

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080221\
Data File : PS016262.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2021 13:37
Operator : DD\AJ
Sample : HSTDICC1500
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 14:01:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080221.M
Quant Title : 8080.M
QLast Update : Mon Aug 02 13:59:52 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

