

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080522\
 Data File : PR055644.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2022 12:48
 Operator : AJ\MA
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 12:58:23 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 12:39:02 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.635	2.905	240.4E6	116.0E6	56.345	57.769
2) SA Decachlor...	9.530	8.124	88977125	92099200	50.042	50.630
Target Compounds						
16) L4 AR-1242-1	4.856	4.013	49487503	29157979	526.157	543.714
17) L4 AR-1242-2	4.878	4.031	70114809	40652647	517.569	543.529
18) L4 AR-1242-3	4.942	4.207	41794326	22158845	524.596	545.890
19) L4 AR-1242-4	5.044	4.299	33386800	20249190	530.115	537.908
20) L4 AR-1242-5	5.832	4.841	29094670	26636623	505.202	538.082



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR080522\
Data File : PR055644.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2022 12:48
Operator : AJ\MA
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 12:58:23 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR080522.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 12:39:02 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.635 min
Delta R.T.: -0.001 min
Response: 240425672
Conc: 56.35 ng/ml

Instrument :
ECD_S
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 2.905 min
Delta R.T.: 0.000 min
Response: 115989337
Conc: 57.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.530 min
Delta R.T.: 0.001 min
Response: 88977125
Conc: 50.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.124 min
Delta R.T.: 0.001 min
Response: 92099200
Conc: 50.63 ng/ml



#16 AR-1242-1


R.T.: 4.856 min
Delta R.T.: 0.000 min
Response: 49487503
Conc: 526.16 ng/ml

Instrument :
ECD_S
ClientSampleId :




#16 AR-1242-1

R.T.: 4.013 min
Delta R.T.: 0.000 min
Response: 29157979
Conc: 543.71 ng/ml




#17 AR-1242-2

R.T.: 4.878 min
Delta R.T.: 0.000 min
Response: 70114809
Conc: 517.57 ng/ml



#17 AR-1242-2


R.T.: 4.031 min
Delta R.T.: 0.000 min
Response: 40652647
Conc: 543.53 ng/ml



#18 AR-1242-3


R.T.: 4.942 min
Delta R.T.: 0.000 min
Response: 41794326
Conc: 524.60 ng/ml

Instrument :
ECD_S
ClientSampleId :




#18 AR-1242-3

R.T.: 4.207 min
Delta R.T.: 0.000 min
Response: 22158845
Conc: 545.89 ng/ml




#19 AR-1242-4

R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 33386800
Conc: 530.11 ng/ml



#19 AR-1242-4


R.T.: 4.299 min
Delta R.T.: 0.000 min
Response: 20249190
Conc: 537.91 ng/ml



#20 AR-1242-5

R.T.: 5.832 min
Delta R.T.: 0.000 min
Response: 29094670
Conc: 505.20 ng/ml

Instrument :
ECD_S
ClientSampleId :



#20 AR-1242-5

R.T.: 4.841 min
Delta R.T.: 0.000 min
Response: 26636623
Conc: 538.08 ng/ml