

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080318\
 Data File : PS001362.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Aug 2018 10:53
 Operator : UA
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 22:46:47 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
 Quant Title : 8080.M
 QLast Update : Sat Jul 21 01:18:58 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	11.634	11.053	137.4E6	132.5E6	740.286	728.402
Target Compounds							
1) T	Dalapon	3.028	2.443	157.0E6	153.2E6	662.132m	672.146
2) T	3,5-DICHL...	10.050	9.201	187.4E6	171.2E6	694.439	671.868m
3) T	4-Nitroph...	11.227	10.200	76507014	66538905	664.767	652.732
5) T	DICAMBA	11.949	11.341	507.0E6	469.7E6	705.170	685.314
6) T	MCPD	12.288	11.565	36389025	30990253	72.400	71.420
7) T	MCPA	12.538	11.941	55512612	45503814	70.391	69.979
8) T	DICHLORPROP	13.141	12.465	129.2E6	118.4E6	707.963	668.292
9) T	2,4-D	13.499	12.936	169.0E6	161.2E6	688.665	679.207
10) T	Pentachlo...	13.888	13.492	2050.4E6	1953.3E6	728.239	709.728
11) T	2,4,5-TP ...	14.758	14.105	804.5E6	760.9E6	716.784	695.865
12) T	2,4,5-T	15.155	14.676	794.2E6	579.2E6	713.710	626.978
13) T	2,4-DB	15.906	15.322	102.0E6	90254985	708.320	677.781
14) T	DINOSEB	17.405	15.753	545.9E6	495.3E6	710.745	696.798
15) T	Picloram	17.165	16.935	997.8E6	927.2E6	631.288	672.624
16) T	DCPA	17.745	16.990	969.8E6	896.0E6	718.333	620.391

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS001362.D
Data File : PS001362.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Aug 2018 10:53
Operator : UA
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 22:46:47 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072118.M
Quant Title : 8080.M
QLast Update : Sat Jul 21 01:18:58 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

