

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080320\
 Data File : PS011606.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Aug 2020 05:21
 Operator : DD\AJ
 Sample : L3507-07MS
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 SU-01-073020MS

Manual Integrations
 APPROVED

Ankita
 8/4/2020 12:01:24 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 04 07:16:38 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
 Quant Title : 8080.M
 QLast Update : Mon Jul 13 16:10:29 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.223	6.916	524.2E6	327.0E6	540.234	580.785
Target Compounds							
1) T	Dalapon	3.036	2.623	491.1E6	426.7E6	408.107	555.898 #
2) T	3,5-DICHL...	6.436	6.067	622.0E6	369.8E6	464.836	440.334
5) T	DICAMBA	7.414	7.084	1950.6E6	1132.1E6	503.272	494.356
6) T	MCP	7.582	7.181	239.4E6	90032165	79.046m	47.184 #
7) T	MCPA	7.731	7.395	279.7E6	203.3E6	65.588	70.531
8) T	DICHLORPROP	8.106	7.710	490.1E6	313.4E6	522.726	520.646
9) T	2,4-D	8.335	8.004	597.1E6	464.7E6	627.839	619.882
10) T	Pentachlo...	8.646	8.433	4302.8E6	2412.3E6	325.510	346.112
11) T	2,4,5-TP ...	9.194	8.790	3049.0E6	3289.7E6	575.749m	1076.037m#
12) T	2,4,5-T	9.480	9.174	2881.3E6	1731.0E6	571.026	570.712
13) T	2,4-DB	10.033	9.680	315.1E6	235.3E6	488.880	550.474
14) T	DINOSEB	11.192	10.013	956.6E6	526.1E6	248.170	245.284
15) T	Picloram	11.013	10.996	2655.3E6	1738.4E6	420.811	397.500
16) T	DCPA	11.457	10.941	3494.9E6	2112.9E6	542.433m	558.695

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080320\
Data File : PS011606.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Aug 2020 05:21
Operator : DD\AJ
Sample : L3507-07MS
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
SU-01-073020MS

Manual Integrations
APPROVED

Ankita
8/4/2020 12:01:24 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 04 07:16:38 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS071320.M
Quant Title : 8080.M
QLast Update : Mon Jul 13 16:10:29 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

