

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080322\
 Data File : PS020146.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Aug 2022 17:25
 Operator : AJ\MA
 Sample : HSTDICC200
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC200

Manual Integrations
 APPROVED

Reviewed By :Ankita Jodhani 08/04/2022
 Supervised By :mohammad ahmed 08/04/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 04 02:04:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 04 01:54:01 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.186	6.763	606.3E6	161.1E6	205.610	196.940m
Target Compounds							
1) T	Dalapon	3.004	2.519	885.7E6	192.5E6	182.941	177.347
2) T	3,5-DICHL...	6.401	5.949	770.6E6	232.7E6	198.101	184.324
3) T	4-Nitroph...	7.035	6.397	199.5E6	83448634	143.660	169.915
5) T	DICAMBA	7.370	6.923	2253.6E6	658.8E6	195.821	184.836
6) T	MCPD	7.533	7.013	147.2E6	45459336	15.872	16.665
7) T	MCPA	7.680	7.223	227.1E6	78586305	17.630	18.668
8) T	DICHLORPROP	8.062	7.544	571.7E6	175.7E6	199.724	187.842
9) T	2,4-D	8.298	7.840	467.6E6	218.3E6	155.182m	186.087
10) T	Pentachlo...	8.595	8.258	8721.8E6	2703.4E6	222.965	211.270
11) T	2,4,5-TP ...	9.147	8.620	3527.9E6	1034.8E6	207.058	193.463
12) T	2,4,5-T	9.436	8.999	3785.5E6	947.0E6	208.551	186.035
13) T	2,4-DB	9.991	9.503	649.2E6	121.1E6	177.899	178.067
14) T	DINOSEB	11.132	9.823	3288.4E6	706.8E6	206.699	186.274
15) T	Picloram	10.969	10.810	6051.1E6	1373.8E6	191.378	182.095
16) T	DCPA	11.401	10.751	5324.3E6	1348.8E6	216.570	203.424

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080322\
Data File : PS020146.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Aug 2022 17:25
Operator : AJ\MA
Sample : HSTDICC200
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC200

Manual Integrations
APPROVED

Reviewed By :Ankita Jodhani 08/04/2022
Supervised By :mohammad ahmed 08/04/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 04 02:04:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
Quant Title : 8080.M
QLast Update : Thu Aug 04 01:54:01 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

