

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080322\
 Data File : PS020149.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Aug 2022 18:38
 Operator : AJ\MA
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 04 01:59:49 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
 Quant Title : 8080.M
 QLast Update : Thu Aug 04 01:54:01 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.181	6.760	2873.6E6	804.7E6	993.571	983.588
Target Compounds							
1) T	Dalapon	3.003	2.518	4307.1E6	976.8E6	897.044	894.740
2) T	3,5-DICHL...	6.397	5.945	3476.9E6	1155.6E6	925.452	917.238
3) T	4-Nitroph...	7.012	6.381	1343.1E6	447.4E6	896.243	893.383
5) T	DICAMBA	7.369	6.923	10498.5E6	3304.2E6	939.965	930.679
6) T	MCPD	7.540	7.021	901.7E6	261.7E6	92.823	92.333
7) T	MCPA	7.689	7.234	1225.2E6	383.8E6	94.291	91.361
8) T	DICHLORPROP	8.060	7.542	2605.6E6	863.1E6	937.550	925.287
9) T	2,4-D	8.290	7.833	2949.5E6	1082.9E6	929.283	925.812
10) T	Pentachlo...	8.595	8.257	35043.9E6	11397.2E6	986.447	948.785
11) T	2,4,5-TP ...	9.145	8.618	15384.1E6	4943.5E6	948.698	942.408
12) T	2,4,5-T	9.430	8.995	16386.0E6	4775.5E6	952.708	942.143
13) T	2,4-DB	9.981	9.499	3489.0E6	648.8E6	938.848	933.707
14) T	DINOSEB	11.131	9.822	14148.2E6	3516.2E6	931.849	929.988
15) T	Picloram	10.957	10.801	29350.9E6	7129.8E6	943.105	941.950
16) T	DCPA	11.402	10.750	21962.8E6	6108.7E6	951.019	955.099

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080322\
Data File : PS020149.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Aug 2022 18:38
Operator : AJ\MA
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 04 01:59:49 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080322.M
Quant Title : 8080.M
QLast Update : Thu Aug 04 01:54:01 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

