

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080323\
 Data File : PS023926.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Aug 2023 19:19
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 08/04/2023
 Supervised By : Ankita Jodhani 08/04/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 20:26:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
 Quant Title : 8080.M
 QLast Update : Thu Jul 20 19:21:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.370	6.804	1446.7E6	491.2E6	729.678	722.260
Target Compounds							
1) T	Dalapon	2.491	2.559	1404.6E6	720.8E6	655.793	650.882
2) T	3,5-DICHL...	5.743	5.984	2100.5E6	723.5E6	680.168	675.625
3) T	4-Nitroph...	6.207	6.415	864.9E6	283.2E6	695.742	668.792
5) T	DICAMBA	6.513	6.968	6349.4E6	2143.4E6	700.854	702.328
8) T	DICHLORPROP	7.102	7.585	1598.7E6	506.7E6	684.317	670.736
9) T	2,4-D	7.299	7.870	1771.2E6	575.5E6	686.569	675.151
10) T	Pentachlo...	7.530	8.297	22955.3E6	7848.6E6	713.891	709.570
11) T	2,4,5-TP ...	8.052	8.655	9116.4E6	2946.5E6	704.960	709.129
12) T	2,4,5-T	8.306	9.027	8495.1E6	2790.9E6	721.347	708.947
13) T	2,4-DB	8.814	9.531	1062.5E6	359.4E6	701.012	724.773
14) T	DINOSEB	9.869	9.859	6329.5E6	2033.7E6	766.006m	735.665
15) T	Picloram	9.692	10.816	10255.7E6	3929.8E6	669.171	647.929
16) T	DCPA	10.134	10.783	10202.0E6	3348.3E6	699.049	704.243

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080323\
Data File : PS023926.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Aug 2023 19:19
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 08/04/2023
Supervised By : Ankita Jodhani 08/04/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 20:26:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
Quant Title : 8080.M
QLast Update : Thu Jul 20 19:21:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

