

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080324\
 Data File : PS027291.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2024 23:25
 Operator : AR\AJ
 Sample : HSTDICC1000
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 03 07:46:18 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080324.M
 Quant Title : 8080.M
 QLast Update : Sat Aug 03 07:31:43 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.188	7.747	2431.0E6	2135.4E6	1019.302	1002.414
Target Compounds							
1) T	Dalapon	2.607	2.696	2944.4E6	3513.2E6	928.928	919.030
2) T	3,5-DICHL...	6.366	6.700	3498.2E6	2945.2E6	947.012	939.845
3) T	4-Nitroph...	6.986	7.275	1420.5E6	1443.0E6	926.676	913.813
5) T	DICAMBA	7.371	7.946	10112.1E6	9535.8E6	959.239	944.299
6) T	MCPD	7.551	8.049	748.7E6	691.3E6	95.401	93.519
7) T	MCPA	7.698	8.293	942.2E6	917.1E6	94.597	92.781
8) T	DICHLORPROP	8.065	8.658	2411.9E6	2301.6E6	956.456	943.814
9) T	2,4-D	8.292	8.987	2662.0E6	2595.8E6	955.970	942.785
10) T	Pentachlo...	8.585	9.515	37031.5E6	36126.5E6	1007.759	986.011
11) T	2,4,5-TP ...	9.157	9.889	14199.1E6	14686.5E6	969.900	961.339
12) T	2,4,5-T	9.446	10.308	14040.6E6	14775.7E6	970.455	964.685
13) T	2,4-DB	10.014	10.873	2061.4E6	1863.5E6	966.269	959.593
14) T	DINOSEB	11.206	11.251	10156.3E6	10102.0E6	960.457	956.262
15) T	Picloram	11.020	12.344	19505.8E6	23455.7E6	969.642	976.926
16) T	DCPA	11.502	12.292	17258.2E6	16522.9E6	980.543	988.749

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080324\
Data File : PS027291.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2024 23:25
Operator : AR\AJ
Sample : HSTDICC1000
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 03 07:46:18 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS080324.M
Quant Title : 8080.M
QLast Update : Sat Aug 03 07:31:43 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

