

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080423\
 Data File : PS023928.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Aug 2023 09:25
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 00:34:31 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
 Quant Title : 8080.M
 QLast Update : Thu Jul 20 19:21:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	6.371	6.803	1009.2E6	331.4E6	509.034	487.302
Target Compounds							
1) T	Dalapon	0.000	2.598	0	67196	N.D.	<MDL
2) T	3,5-DICHL...	5.709	0.000	1874459	0	<MDL	N.D. #
3) T	4-Nitroph...	6.212	0.000	8252914	0	6.638	N.D. #
5) T	DICAMBA	6.512	6.966	757191	616903	<MDL	<MDL #
8) T	DICHLORPROP	7.104	7.601	-26040	2463495	N.D.	3.261
9) T	2,4-D	7.267	7.816f	3628588	884834	1.407	1.038 #
10) T	Pentachlo...	7.529	8.299	4435435	605361	<MDL	<MDL #
11) T	2,4,5-TP ...	8.051	8.655	2393757	822450	<MDL	<MDL #
12) T	2,4,5-T	8.310	9.028	5687962	1055455	<MDL	<MDL #
13) T	2,4-DB	8.760f	0.000	1328143	0	<MDL	N.D. #
14) T	DINOSEB	9.870	9.860	1902942	503600	<MDL	<MDL #
15) T	Picloram	9.697	10.817	103.2E6	36586395	6.731	6.032
16) T	DCPA	10.133	10.817	2344443	36586395	<MDL	7.695 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS080423\
Data File : PS023928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Aug 2023 09:25
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 00:34:31 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS072023.M
Quant Title : 8080.M
QLast Update : Thu Jul 20 19:21:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : Rtx-CLPesticides Signal #2 Phase: Rtx-CLPesticides2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm






#3 4-Nitrophenol


R.T.: 6.212 min
Delta R.T.: 0.002 min
Response: 8252914
Conc: 6.64 ng/ml

Instrument :
ECD_S
ClientSampleId :




#3 4-Nitrophenol

R.T.: 0.000 min
Exp R.T. : 6.419 min
Response: 0
Conc: N.D.




#4 2,4-DCAA

R.T.: 6.371 min
Delta R.T.: -0.002 min
Response: 1009230266
Conc: 509.03 ng/ml



#4 2,4-DCAA


R.T.: 6.803 min
Delta R.T.: -0.004 min
Response: 331413040
Conc: 487.30 ng/ml



#8 DICHLORPROP


R.T.: 7.104 min
Delta R.T.: 0.000 min
Response: -26040
Conc: N.D.

Instrument :
ECD_S
ClientSampleId :




#8 DICHLORPROP

R.T.: 7.601 min
Delta R.T.: 0.012 min
Response: 2463495
Conc: 3.26 ng/ml




#9 2,4-D

R.T.: 7.267 min
Delta R.T.: -0.035 min
Response: 3628588
Conc: 1.41 ng/ml



#9 2,4-D


R.T.: 7.816 min
Delta R.T.: -0.059 min
Response: 884834
Conc: 1.04 ng/ml



#15 Picloram


R.T.: 9.697 min
Delta R.T.: 0.000 min
Response: 103156585
Conc: 6.73 ng/ml

Instrument :
ECD_S
ClientSampleId :




#15 Picloram

R.T.: 10.817 min
Delta R.T.: -0.006 min
Response: 36586395
Conc: 6.03 ng/ml



#16 DCPA

R.T.: 10.133 min
Delta R.T.: -0.005 min
Response: 2344443
Conc: N.D.



#16 DCPA

R.T.: 10.817 min
Delta R.T.: 0.028 min
Response: 36586395
Conc: 7.70 ng/ml